Adding Brand

Design, Branding and Additive Manufacture in the production of tangible prosthetic products.

© MATT MCGOWAN 2020

Adding Brand

Design, Branding and Additive Manufacture in the production of tangible prosthetic products.

Matt McGowan

Victoria University of Wellington
School of Design 2020

Supervised by Bernard Guy and Simon Fraser

Figure 1. Images of a hand fabricated, lower-limb prosthetic.

Abstract

Design, Branding and Additive Manufacture in the production of tangible prosthetic products.

For the New Zealand Artificial Limbs Service, (NZALS) prosthetic design has skipped the mechanisation and mass production paradigm seen in the automation of consumer production. This industry predominantly uses traditional hand fabrication methods to produce prosthetics as a one-off appendage. This research asks; how can design communicate the possibilities of Additive Manufacture?

This research addresses the creation of branded designed products for the NZALS, and as a result, exposes the predominantly service based industry of the NZALS to a product focused methodology through traditional industrial design practices. This has been achievable by investigating emerging platforms of manufacturing in both Digital and Additive Manufacturing (3D printing), with the development of the designs in this research focused on brand, client and company identity. This focus addresses the integration of an Industry 4.0 model in favour of the amputee client, and realises future outlooks of prosthetic production envisioned by the NZLAS.

Design research in this thesis has seen the creation of two prosthetic products. Firstly, a below-the- knee Prosthetic Fairing (Easycover), and secondly, a fully 3D printed below-the- knee prosthetic (Easylimb). The research undertaken shows the importance of creating tangible and readymade products to allow the NZALS, it's staff and amputee clientele, to understand the benefits of design, branding, and emerging platforms of manufacture in the production of prosthetic diversity.

'We found patients were more likely to respond and contribute to design development where ideas appear to be complete, tangible, readymade and resemble the expectation of an off the shelf product'

(Aaron Johnston 2019 NZALS prosthetic technician)

Acknowledgements

To Bernard and Simon, thank you for your expertise and personal evaluation throughout this process. The knowledge I have gained through your guidance, and through this process, has been invaluable in developing my design skills for the future.

To Aaron Johnston, thank you for your continued input throughout this research. Your professional evaluation and expertise has been imperative in developing ideas from paper into tangible, physical concepts.

To Sean Grey, thank you for your enthusiasm and financial support throughout this research. Your continued validation of this process has provided me with the motivation to finish this research to a high standard.

Thank you to the NZALS for their continued financial support throughout this research, and for the welcoming access to the NZALS center in Newtown.

Thank you to Hamish Trolove at Callaghan Innovation for processing prototyping prints and providing information about file management.

To my parents, Pete and Dianne, your support and assurance has supported me through countless times over the last 5 years. I strive to uphold your hardworking and driven attitudes.

Contents

	-Technology and the NZALS	29
	-Research opportunities	30
-Introduction	-Past VUW interactions	33-37
-Research question	2. Precedent analysis	38
-Terminology	-Allbirds	41
1. Background research	-Uynq and Alleles	43
-The NZLAS	-Research through design and industrial design precedent	45
-Traditional and current prosthetic fabrication	-Eyewear/ prescription glasses	47
-NZALS Aims and Objectives	-Practicum as research	49
-NZLAS as a service	3. Methodologies	50
-Industry 4.0	-Research Methodologies	53
-Branding within a product	-Methodology foundation	57
-Importance of branding	-Role of the designer	58
-Emerging platforms of manufacturing	-Research through design	59

-Industrial design process	61-64	-Easylimb final modeling	242-281
-Interaction process	65-67	-Easylimb product visualisation	282-317
-Software	68	-Easylimb final product	318-335
-Digital workflow	70	-Easylimb contextual website	336-355
-3D printing processes	72	7. Discussion	356-361
-3D printing materials	74	-Limitations and future speculations	362-371
-External 3D printing	77	8. Conclusion	373
4. Application	78	9. Sources	374
- Logo design	80-89	10. Figures list	376
5. Easycover	90-143		
6. Easylimb	144-165		
-Easylimb initial modeling	166-173		
-Easylimb socket modeling	174-183		
-Easylimb sole modeling	184-241		

Introduction

The NZLAS operates within a service based industry, in that the services offered by the company address the essential clinical assessment, assembly of components, and fitting of prosthesis. The services offered by the company predominantly still use traditional, hand-fabrication methods to produce prosthetics.

These fitting and fabrication techniques rely on a skilled prosthetic technician to create prosthetic limb sockets (for lower-limb prosthesis) from scratch. Fibreglass is used and then an assortment of expensive alloy and composite components are added to create the prosthesis.

The linear service process of an amputee visiting the NZALS for prosthetic fabrication and adjustments creates opportunities for greater business innovation within the company. These opportunities include, adopting emerging platforms of Digital Manufacture for prosthetic fabrication, along with a product focused methodology for the creation of prosthetic products.

How this could be communicated to the service-based industry of the NZLAS is the focus of this research.

“3D printing has the potential to significantly evolve the service experience of patients.” (NZALS 2017)

These opportunities also extend to creating product diversity through the creation of considered product ranges. Currently, the 3D printing paradigm seen within the prosthetic industry is used typically to create bespoke one-off appendages or designs.

This research aims to oppose the established precedent that Additive Manufacturing is used solely to create bespoke one-off prosthetics, and shows how 3D printing can be applicable to the prosthetic industry, by creating commercial products that are understandable and applicable to a wide range of amputees.

AM technology and processes have transformed conventional product design. Although 3D printing has been used for rapid prototyping, its use as a primary production method is imminent in the future of mass and tailored manufacturing.

The first prompts for adopting AM into the NZALS have been identified as the production of 3D printed, Multi-jet fusion prosthetic sockets for selected amputees. These first tangible applications illustrate how AM can benefit the amputee though a more accurate prosthetic fit. It also shows efficiencies within a digital and reproducible database, where diagnostic data from the amputee can be stored and used when needed.

This shows the NZLAS's ability to adopt innovation and technology via Industry 4.0 production to improve it's prosthetic production, and overall patient care.

This research prompts how design can be used as a process to integrate these production methods into prosthetic production and to illustrate product diversity. This includes using traditional industrial design processes though prototyping and extensive iteration, using Additive Manufacture as a prototyping tool, and production process for final concept validation.

This research communicates two physical outputs in the form of tangible, branded products. Firstly, a cosmetic Prosthetic Fairing (the Easycover) and secondly, a full-length below- the knee prosthetic (the Easylimb). It also provides examples of how these products can be communicated in a product range through an online purchasing platform, where the shopping experience is normalised for the amputee.

Additive Manufacture is used as the singular production process to create these tangible product concepts.

The Easylimb and Easycover aim to communicate visually understandable and credible manifestations of this research, with the outputs of this research aiming to provide a portfolio for experience based interactions between the NZLAS, it's workforce and the amputee client.

—

Proposed research question:
How can design communicate the possibilities of Additive Manufacture?

Terminology

1. NZALS: The New Zealand Artificial Limbs Service.

2. AM: Additive Manufacture. (3D printing)

3. Traditional/existing prosthetic: Prosthesis that consists of a custom-made fibreglass socket, a titanium pylon and carbon fibre or rubber composite foot.

4. Prosthetic Fairing: Cosmetic cover that fits over the socket, pylon, and other prosthetic components.

5. NZALS technician: Prosthetic expert who undertakes prosthetic fitting and fabrication. These professionals closely interact with prosthetic clinicians in the fabrication of a prosthetic.

6. Callaghan Innovation: New Zealand based organisation that has the task of making New Zealand business more innovative.

7. Materialise: An external 3D printing company that processes 3D files into 3D printed prototypes.

8. Secondary limb: Complementary daily activity limb used in rotation with a traditional prosthetic.

9. Aesthetics: Concerned with visual appearance or beauty.

10. Tangible: Capable of being touched; physical rather than conceptual.

11. Industry 4.0: Transformation of traditional manufacturing and industrial practices combined with the latest smart technology.

12. Emerging platforms of manufacture: 3D printing and digital systems i.e. CAD, 3D scanning.

13: Easycover: Prosthetic Fairing created in this research.

14: Easylimb: Secondary, 3D printed lower-limb created in this research.

15: Sketch models/Prototypes: Preliminary, physical 3D models used to develop ideas into resolved products.

16: Iteration: The repetition of concepts, each building on the one preceding.

17: CAD: Computer Aided Design/ Digital workspace where 3D models can be created.

18: Polysurface/Nurbs surface: A type of surface created in a CAD modeling software used to define a solid object or 3D model. (Often defined by a curve cage)

19: 3D mesh: A collection of vertices, edges and faces that defines the shape of an object or 3D model.

20: Curve cage: Set of lines used to define the surface of a digital 3D model in a CAD software.

21: Rhino: CAD modeling software used throughout this research.

BACKGROUND RESEARCH

This section provides a context for the research. Background research was undertaken on areas relevant to the NZLAS. This included exploring the NZALS and it's position as a service industry, Additive Manufacturing, and the integration of Industry 4.0 into the company's prosthetic production. Other relevant information was gained from existing literature to provide a framework for the research.

Framing the research from existing literature and gaining information about the NZLAS, was important to align the products created in this research to NZLAS goals and future outlooks, as well as producing a valid research portfolio.

The NZLAS

History

After WW1 and WW2, the creation of prosthetics began to be significant, due to soldiers returning back from the battlefields of Europe and the Pacific to New Zealand.

The advancements in sanitation, anaesthesia and surgery meant that increasing numbers of amputees were surviving the injuries they sustained. The global mass production of prosthetics, as a consequence of war, enabled returning soldiers in NZ to be reintegrated immediately back into working life, via the limbs fitting services provided. For NZ, one of the main motivations for this reintegration, was to increase economic production within primary industries, as this had been affected as a consequence of a diminished workforce.

Although there has been many advancements in the technology integrated into prosthetic components via the third industrial revolution (integration of microprocessors, new alloy metals and plastics and the mass production of components), the fitting and fabrication techniques used today are very similar to the processes used back in Post-war prosthetic limbs services.

For the NZALS, prosthetic design has skipped the mechanisation and mass production paradigm seen in the automation of production. The services offered by the company, predominantly still use traditional handcrafted fabrication methods to produce prosthetics (Shuxian et al. 2004), (NZALS 2017)

Figure 2. Hand fabrication of a prosthetic at the NZLAS

Traditional and current prosthetic fabrication

Traditional prosthetic fabrication for sockets, via mould creation, as explained by Herbert (2005), results in a time-consuming production process, that involves multiple components and steps in order to create one single prosthesis. Along with this, traditional and current prosthetic fabrication can produce a prosthetic socket that "requires constant reiteration, including reshaping" (Herbert 2005). Current fabrication may also be seen as an inefficient production method due to the subtractive fabrication methods using plaster and moulding techniques.

AM allows for the creation of prosthetics that produce significantly less material waste in comparison to traditional fabrication, due to the additive production method. (Bhatia, Sharma 2014).

In "The leg shop" Evans (2019), explains how the traditional fabrication of prosthetics via the NZLAS workshop, requires prosthetic technicians to use the following tools and skill sets in order to produce sockets that function appropriately for amputees. "Prosthetists use sewing needles, band saws and grinding machines on a daily basis, manipulating leather, silicon, resin, and plaster of Paris with subtlety and perseverance"

Due to time and labour being excessive in prosthetic production, amputees are often provided with a single prosthetic which is a combination of a hand fabricated fibreglass socket, and expensive, imported alloy and composite hardware.

Figure 3. Hand fabricated below the- knee- prosthetic created at the NZALS

NZALS Aims and Objectives

Current objectives outlined by the NZALS within their annual report (2018) and statement of intent (2017-2021) include, (among many) the following selected objectives for the future outputs and goals for the company. These following goals align to the ideas expressed within this research.

These future goals state:

1. "To Explore new technology and its potential applications to limb making"
2. "Collaborating with other technology developers and users, including staff exchanges, career development, and better technology solutions for patients"
3. "Investing in research that will benefit our patients"
4. "Providing a collaborative pathway to transfer technology from an idea to delivery and outcome"

The NZALS statement of intent (2017-2021), explains how the advancing technology within prosthetic development that is accessible to patients, is constantly changing. AM technology is described as a "disruptive technology advancement" to traditional services models under the "Technology and R&D factors" section of their 2017-2021 statement of intent (NZALS 2017-2021).

An Industry 4.0 model, as well as providing an example of tangible AM products, could help change this concept of AM being a disruptive technology into, as mentioned, a primary production method for prosthetic fabrication, leading to greater business innovation.

As stated, this research aims to overcome the issues mentioned by the NZALS and expressed by past interactions with the NZALS and VUW, by moving away from the primary initiative of researching 3D printing as a prototyping method; "In the early stage of development, new technologies often lack the necessary evidence to support funding and prescription decisions." (NZALS 2017-2021).

"AM technologies need to be developed in a considered way that balances the benefits and risks of AM technology to ensure the appropriate implementation of 3D printing technology into service delivery" (NZALS 2017-2021).

This research aims to overcome this premise expressed by the company, by communicating to the service based industry of the NZLAS, that design and AM can be used to create product diversity through industrial design. And, that these products are understandable and appropriate to the client and company.

NZALS as a service

The defining term of "service" within the name "New Zealand Artificial Limbs Service" defines the operations of the company.

This service is the essential clinical evaluation, assemblage and fitting of providing amputees with a functional prosthetic.

Amputees will constantly be a part of the NZALS as the refitting and reshaping of an individual's prosthetic will entail retuning visits to the nearest NZALS centre. (Evans 2019).

This linear service process of an amputee visiting the NZALS for prosthetic fabrication, and adjustments, leaves open a space to offer more experience based interactions between amputees and the NZALS. This potentially could build brand image and equity.

AM technology, and the creation of branded AM products, provides a physical interaction for the amputee. This interaction may illustrate that the NZALS is a brand name focused on technology and innovation and that the amputee can be a part of these experiences and the greater innovation of the brand.

Industry 4.0

Callaghan Innovation (2019) describes Industry 4.0 as “the new era of ‘Smart Manufacturing’”. This modern industry model, employs that companies focused on innovation, seek to incorporate a range of digital technologies within their production (or services) and greater business strategies.

These digital technologies are incorporated to “enhance the performance, output, monitoring and control of their manufacturing processes, giving them a significant edge over their competitors” (Callaghan Innovation 2019)

In “Industry 4.0”, Lasi, et al (2014) suggests that the advanced digitalization within production, and the combination of modern manufacturing techniques, seems to result in a “new fundamental paradigm shift in industrial production.”

Importantly, the term Industry 4.0 argues for a new industrial revolution (Gunal 2019). This means modern manufacturing techniques (such as AM) are combined with digital network-ing, to create unified manufacturing systems that are able to communicate and analyse information to provide feedback into the physical output of production.

This research seeks out to theoretically apply an Industry 4.0 model, where Additive and Digital manufacture are used in the design and manufacturing of products/concepts. CAD modelling, 3D printing and 3D scanning are all parts of these digital systems, and are used throughout this investigation.

Additionally, Digital Manufacturing and systems within Industry 4.0 open the possibilities of creating a digital database for the NZLAS, where amputee’s diagnostic data can be stored and retrieved on demand.

Parametric capabilities within CAD programs allow a set product to be digitally altered and fitted to an individual’s scan. Other benefits of this personal tailoring, are that the company can maintain the economic conditions of mass production, and attend to remote customers. Lasi et al (2014) describes this concept as “Individualization on demand”.

A more applicable interpretation may be ‘product variation within a product range’ whereby utilising design to address individual style and taste, the NZLAS can create products that cater to a large percentage of amputees whilst fitting each individual.

Branding within a product

In 'Brand New', Pavitt (2000) discusses how “brand values are literally built into products; it is not a process of simply styling but development and innovation”.

The NZLAS annual report for 2018 states “The NZALS workforce is able to develop adapt and adopt innovation and technology, to continually improve the service to and outcomes for amputees (NZALS 2018). Therefore, the creation of tangible products may be an embodiment of this adoption of innovation, and can be expressed as a brand value for the company.

Deciding how to integrate branding into a product in the whole stage of the design process is important to define the brand through the product. But also, this is important to create a connection between the designers themselves, and how they want the product to be experienced by the consumer to illustrate brand values.

Emerging products express the brand values that consumers like to be associated with, such as being innovative, and the ability to cater for individual's personal taste and style. This needs to be addressed if product diversity is to be realised in the creation of an AM product range.

Importance of branding

To enable the NZALS to visualize a product range, there needs to be an overarching brand name and product names that can be marketable. For the NZLAS, the goal is to establish their brand as market innovators.

“Brands that are market leaders have often been established during a major shift in the market. In many cases they have been at the forefront of innovation during the shift” (Hestad, 2013).

This can be a direct relationship between the use of Additive and Digital Manufacture within prosthetic design. Building a brand that shows it is an industry innovator, also provides a competitive advantage over existing companies' business models.

In “branding and product design” Hestad (2013), explains how designs and branding should be embedded in what the design process and that the integration of these two disciplines should be driven by company values.

This is also elaborated on by Mycroft (2011), who explains how product designers have changed, to not only design, but design brand into their products. “Branding and product design have typically been seen as two separate activities, while today designers will have to incorporate brand awareness into the development process” (Mycroft 2011)

Developing brand awareness is a process that involves creating a broad, but strong vision of what the company and its products should express to the consumer. (Hestad 2013)

Establishing a strong vision of how the branded product should appear early in the design process is one significant way to work towards building brand awareness for the company, as well as solidifying branding within the product itself for resolved designs. (Hestad 2013).

Company's that express strong brand identity through their products, and product experiences, often initiate change to product development and technology, to stay up to-date and maintain their position as market leaders. New technologies, such as 3D printing, are envisioned and realised as value propositions by the NZALS, but are yet to be fully implemented and streamlined into their prosthetic production.

This research will begin to show how Additive Manufacture can link with brand and designed form, proposing an innovative paradigm shift in the companies' prosthetic development.

Emerging platforms of manufacturing

Additive Manufacturing techniques and technologies allow for the creation of complex geometric shapes that can be mass-customized due to fewer limitations than in traditional manufacturing techniques.

Design concepts, that otherwise would be difficult to produce through traditional manufacturing or prototyping methods, can be prototyped easily through the use of Additive Manufacture, and can be translated into products easily through direct Digital Manufacturing.

The additively layered approach used in 3D printing, enables the merging of multiple components into a single piece, which removes the requirement for complex assembly operations or the creation of multiple components (Macdonald, Wicker 2016).

The tool-less process of 3D printing permits the creation of functional and visual forms that allow greater diversity in 3D form, leading to aesthetic choice, rapid reproduction, and economies in both logistics, material and time. (Toma 2018).

The addition of 3D printing technologies and CAD/CAM software, also allows for the creation and fitting of prosthetic sockets exponentially more efficiently, when labour, material, components, and processes are considered.

The relationship between 3D printing technologies and other modern digital prototyping methods, such as 3D scanning and CAD modelling, may allow for the development of prosthetics which have proper fit and comfort, as well as improving function.

Tsur (2018) explains, "The critical relationship between the accurate fit of a prosthesis and amputee comfort and function constitutes the foundation of a successful prosthetic."

Technology and the NZLAS

AM technology and processes have transformed conventional product design, and although 3D printing has been used for rapid prototyping, its use as a primary production method is imminent in the future of mass and bespoke manufacturing.

One reason for this, is because Digital and AM technology allows for the simulation of production, before physical outputs are produced, increasing efficiency and productivity. (Gunal 2019)

AM technology is realised by the NZALS as a viable future technology to advance the development of its prosthetic design. Hence, the collaborative efforts with VUW to produce and experiment with AM technology as a prototyping method. (NZLAS 2017-2021). This technology, however, has yet to be fully integrated into the production of its prosthetics.

The introduction to AM and smart 3D technologies, such as 3D scanning, contributes to the shift away from traditional fabrication for the NZALS through their research and development. (NZLAS 2017-2021).

The first prompts for the functional integration of AM into the NZALS has been identified as a small run batch of Multi- jet fusion, 3D printed, below-the-knee sockets for selected amputees.

Amputee's limbs are able to be scanned into the computer, and a bespoke socket is able to be created for them using 3D software such as Mesh-Mixer. The 3D CAD model is able to be transferred through other 3D software such as Rhino 3D, to add functional componentry, with the end 3D file being able to be manufactured via multi-jet fusion 3D printing at external 3D printing company Shapeways.

This 3D process is considerably more time efficient than current fabrication, and due to AM as a production method, creates minimal excess material in the process. This illustrates the shift from experimental prototyping with AM (mainly via VUW), to the integration of AM into a tangible application

These first tangible applications illustrate how AM can benefit the amputee due to a more accurate prosthetic fit. But, more relevant to this, this research shows how AM can improve the NZALS in terms of its productivity, and also the company's ability to adopt innovation and technology via industry 4.0 methods. (NZLAS 2018)

This research aims to extend this adoption of AM in the company by communicating that design and AM production can lead to the creation of products and product diversity.

Figure 4. Research opportunities – configuring NZLAS's goals with emerging platforms of manufacture for the creation of tangible, readymade products

Past VUW interactions

Victoria University has worked with the NZALS since 2015, primarily researching 3D printing as a prototyping method.

Previous projects have added to the established research for the NZALS, such as the swim limb designed by Stuart Baynes (Baynes, Miller, Guy, 2016). However, these past instances of research have led to little in the way of a resolved product ready for an 'off-the-shelf' application. (See Figure 5)

The NZALS statement of intent (2017-2021) states that "AM technologies need to be developed in a considered way that balances the benefits and risks of AM technology to ensure the appropriate implementation of 3D printing technology into service delivery".

In this context, the creation of resolved designs that resemble 'off-the-shelf' products may allow the NZALS, its staff and amputee clients to understand the benefits of design and AM technology. (see Figure 6)

Figure 5. The process of past student and NZLAS interactions – Leading to 3D printing as a prototyping tool - Precedent for approaching the research

Figure 6. The interaction process and approach for this research – configured after reviewing previous student and NZALS interactions

PRECEDENT ANALYSIS

This section aims to review relevant design and business precedents, to inform the design process. It also provides a context for approaching branding, visual language and the methods and processes used to design a product.

The images included have been taken from screenshots via the precedent companies' websites and public domain documents. This is important to reinforce these products/companies as marketable and commercial entities.

Figure 7. Screenshots taken from the Allbirds website containing the company's products (Allbirds 2020)

Allbirds

Relevant to this research was finding a company precedent that showed a distinct visual language expressed throughout a product range. Allbirds was chosen as a precedent company because of its increasing popularity and success in the footwear market. In a design context, Allbirds' subtly designed products, enable the brand to appeal to a large percentage of the footwear market. This is complemented with the product range offering a large selection of design variation in terms of different models and colour combinations.

Allbirds also has a successful marketing campaign that expresses its products through several purchasing platforms, including physical stores and online websites. Marketing also expresses Allbirds as a brand that addresses customer's lifestyles, personal style, and the quality that is desired by them. If a product range is to be designed for the NZLAS, it is important to realise that the amputee client must be shown a set range of products with visual variation to appeal to their different customer tastes.

- Visual continuity within the product range reflects the brand's ethos. Keeping overall designed form subtle, allows material qualities to guide the visual language (i.e. woollen construction throughout the product). In this respect, the product can remain appealing to a large variety of consumers as the overall visual form of the product is relatively neutral.

- Accessible pricing. If a designed prosthetic range is to be successful it should be priced so that it is accessible to 90% of amputee clients. Allbirds products' retail under \$200NZD which is reasonable and accessible to a large percentage of the population.

Other Key-features of the company's brand that are relevant to this research include...

- A focus on using innovative materials that are integrated into the brand's designs, but also define the brands standout position in the sneaker market. This can be related to this research, as AM will create a point of difference in prosthetic production for the NZLAS.

Figure 8. Screenshots taken from the Uynq and Alleles websites with their Prosthetic Fairing designs (Uynq 2020, Alleles 2020)

Uynq and Alleles

Uynq and Alleles are existing companies that have established brands through the creation of "ready- to- wear" Prosthetic Fairings. "We offer amputees revolutionary personalization in terms of colour and design". "Our covers reflect our customer's prior shape, and the available design and colour options reflect our client's style." (Uynq 2020).

Discussions with NZALS technician Aaron Johnston, revealed that the designs of such products, were sometimes difficult to fit to amputee's prosthetic, if the designs were "one- size fits- all" This was particularly the case with Alleles Prosthetic Fairings.

This created a design opportunity in the first stages of this research, which saw the development of an easy-use, Prosthetic Fairing (Easycover).

The main critique of these brands from a design perspective, is that they have minimal visual coherency throughout the product range.

Each design is significantly different from that of the last. Furthermore, branding is seldom included on the product and therefore minimises brand awareness.

Offering choice within a product range is essential; however, if individual designs appear too bespoke, it may minimise the opportunity for the brand to develop its own design language.

This in turn may minimise the brand/ consumer relationship as consumers have a minimal brand identity to align with.

This reiterates what Pavitt (2000) discusses concerning the importance, of not just styling products, but building brand values into products.

These companies were an important precedent to leverage off, and determine where a point of difference could be made for the design of a similar product (costmetic Prosthetic Fairing).

Figure 9. Page from "TEXTBOOK", digital thesis (Abloh 2017)

Research through design and industrial design precedent

Virgil Abloh, Nike collaboration

Fashion designer Virgil Abloh is recognized for his work within his own fashion brand OFF-WHITE. A collaborative project with Nike in 2017 saw the re-design of several of Nike's most iconic footwear silhouettes throughout the company's History.

Although not an industrial designer, the design process of re-working these iconic designs was documented in a workbook dubbed "TEXTBOOK". The documentation process, along with images displayed in the document, show extensive iterative practice and a research through design process, aimed at creating a new product. "The form and shape and design vernacular is so unique that it created an entirely new form and iconic silhouette" (Abloh 2017)

This "textbook" was important in realising the different methods of documentation used in a professional industrial design setting i.e.: Annotating models, sketching over printed out images and collating visual language to use as references.

"The result is a collage of stories and techniques, releasing the energy built by each Nike icon" (Warnett 2017)

Abloh's process demonstrates what Martin and Hanington (2012) mention when describing the components of research through design "...Including materials research, development work, and the critical act of recording and communicating the steps, experiments, and iterations of design".

As a precedent this aided this research by creating a foundation of how to document and undergo the iterative process.

Figure 10. Screenshots from the Specsavers and Rayban websites

Eyewear/ Prescription Glasses

A good analogy of streamlining prosthetic products into the NZLAS, and how they are exposed to the amputee consumer, is eyewear. There are several components and people who are responsible for an individual receiving a pair of prescription Glasses.

These components can be directly compared to the roles involved in this research, and the envisioned implementation of AM products into the NZALS.

For example:

- Lenses= Prosthetic

The essential component/ product which needs to be purchased in order for the person to see/walk.

- Optometrist= NZLAS technician

The professional who has the technical knowledge to ensure the client has the best possible experience with the product, (tailors to individual specifications). This professional is responsible for clinical knowledge.

- Frames/ Eyewear designer= designed prosthetic/designer

Companies such as Ray ban, specialise in designing frames, which are available in a collection of styles to suit a range of individual's personalities and aesthetic choices.

These frames can be brought to an optometrist to have lenses fitted, which suit the individual's specifications. The designer's role is to maximise the aesthetic and functional qualities of the product, to ensure it is received by the customer as a well resolved item.

- Store= NZAL

Companies such as Specsavers, or again Ray ban, offer physical stores or online stores so customers can access their products or services. Resolved and finished products can be purchased and then tailored to individual specifications. It is envisioned that the products designed in this research can be purchased in such a way.

Figure 11. Prosthetic Fairing designs created by the designer through the practicum with the NZLAS

Practicum as applied research

Having a practicum course running in parallel with this research, during the first three months, meant learning design methods that were transferable into the research, and acted as a trialing period for the methodologies explained further in the research.

Significant transferrable skills and methods included the following:

1. Information gained about the NZALS as a company from informal discussion with NZALS technician Aaron Johnston
2. Learning the CAD software that has been used throughout this research
3. Being informed about prosthetic fitting and clinical aspects of prosthetic componentry and fitting (on an informal basis with NZALS technician's)
4. Information gained about the NZALS and its use of AM
5. NZALS initiatives for the future, via CEO Sean Grey

This section outlines the approach and methods used to undertake this research, and why they were suitable for this application. Traditional industrial design methods, and research through design, were applied throughout the design process/ research from ideation phases to product refinement.

‘Where the design process normally functions to provide concepts which help to inform focus groups and capture feedback from trials, we found that this process was cumbersome and confusing for patients with limited understanding and time to interact’

(Aaron Johnston 2019 NZALS prosthetic technician)

Research Methodologies

The methodologies section of the research is broken down into the following components.

1. Methodology foundation
2. Role of the designer
3. Research through design
4. Industrial Design process
5. Interaction process
6. Software
7. 3D printing process
8. 3D printing materials
9. External 3D printing

“In the early stage of development, new technologies often lack the necessary evidence to support funding and prescription decisions.”
(NZALS 2017-2021statement of intent)

“The designer's contribution to knowledge more typically consists of the development of a prototype (or artefact) that could be mistaken for a 'product', and that plays a central role in knowledge-generation" (Stappers & Giaccardi 2013)

1. Methodology foundation

Within this research, the aim is not to design directly for (and with) the amputee for the creation of a one-off appendage.

Looking at prosthetic product design from a bespoke perspective, greatly limits the opportunity for the creation of a coherent product range, as designs are only suitable to one individual.

The creation of AM products needs to be visually understandable and credible to a range of clients and consumers.

Amputees and NZLAS staff would benefit from being shown a range of designs that are coherent as a collective, but offer varied design language. By creating a product range, it may show the NZALS brand addresses customer's wants, personalities and style.

It also may communicate products as resolved finished items, that are not one-off concepts, intended for a single amputee.

In addition to this, there is the opportunity to normalise the shopping experience of purchasing prosthetic products.

Streamlining Digital and Additive Manufacture into prosthetic production may offer product choice, and with it, the opportunity to market AM products in a more commercial setting.

The adoption of AM to the NZALS is clear. However, there is a need to communicate the possibilities of a well resolved product range, to fully realise the implementation of AM production into the NZALS. Thus the challenge for the research.

2. Role of the designer

When designing a new product from scratch using Digital and Additive Manufacture, the designer must undertake challenging iterative steps in order to develop and refine a design.

Essential to this process, is continued and varied physical prototyping, to validate ideas and digital models created in 3D CAD modeling software.

Throughout this research, the designer's role has been to consciously consider subtle aesthetic and functional refinements within the design process. This includes, extensively reworking, and refining designs for the motivation of providing the NZALS and the amputee client with the best possible envisioned product experience.

This includes, hundreds of hours mastering CAD modeling programs, and countless more hours solving technical problems relating to the AM process. This iterative process is critical if products are to appear tangible, and appeal to a comprehensive range of individuals.

To the consumer, the designer's control and management of 3D form is often unseen. However it is these calculated decisions by the designer, that ultimately make the design visually and functionally desirable.

3. Research through design

Designing a product from scratch using emerging platforms of manufacturing, calls for the design process to involve extensive iteration and lateral thinking. This is in order to refine ideas from concepts into tangible objects.

Industrial design methods such as sketching, visual mapping and physical prototyping, facilitate knowledge in the design process and allow for development.

Research through design is a methodology that can be defined as an investigation that revolves around the creation of a product, service or system.

The knowledge gained through the design process can be embedded in the resulting artefact (product) that is created through the research. (Zimmerman, Stolterman and Forlizzi, 2010).

In unison with this, Martin, Hanington (2012) explains how Research through design is created by the design process itself ... "including materials research, development work, and the critical act of recording and communicating the steps, experiments, and iterations of design".

The resulting product, in this case Designed prosthetic products, can be seen as an artefact which the NZLAS can develop ideas/concepts further.

Product artefacts inform continued research or product development, that can help the designer decide what is relevant in the new design space. (Fraying 1993).

The creation of a physical product concept, or artefact, is essential to visualising and communicating the benefits of AM to the NZLAS and amputee client.

“The final decisions on a products form normally take place in the last phases of the design process, but it is important to realize that before this, the designer has already been reflecting and making decisions that have a fundamental influence on the form.” (Tovsky,1989)

“Drawings and 3D models are the physical manifestation of visual thinking, and provide a language for handling design ideas”

“The actual process of creating design ideas goes on in the mind, and the drawings and 3D models are attempts to reproduce the designer’s mental images” (Tovsky,1989)

4. Industrial Design Process (research through design)

The design process used throughout this research has been informed by industrial design methods, which included extensive iteration. This research has included iterative prototyping and iterative CAD modeling, showing the development of visual and functional form, as well as design ideas.

Figure 12. Iterative design process/workflow use throughout this research

Idea and concept generation

Concept generation involves generating design ideas based on technician expertise and precedent analysis. Industrial design methods, such as creating mood boards, and documenting workflow by compiling work in sketchbooks, have been used to keep track of design decisions and help formulate ideas.

“Visual mapping can be an important tool in enabling products to succeed, as it allows the design team to establish what else is out there, and whether the product is different from other products in the current marketplace”(Hestad 2013).

Sketching and annotations within these workbooks have been used as visual and mental notes.

To frame the research, an early stage design goal (as mentioned in the background research) was to compile a range of precedents images (through workbooks and online platforms such as Pintrest), where leading companies used strong design language to build their brand identity.

CAD Modeling

Being able to visualize an idea in a 3D space is critical to the success of the design process. This means mastering 3D form with competency, to achieve clear visual communication, and create working prototypes. This research demands extensive iteration of complex models within CAD modeling programs such as Rhino 3D and Netfabb.

As Rhino and Netfabb were learned in parallel with this research, it was crucial to get up to speed with the programs in order to create models that would appear tangible and could be produced with Additive Manufacture.

Digital visualization

A range of visualization tools used within Rhino and Keyshot have been used throughout this research, to analyse the visual qualities of CAD models. This is to identify where subtle changes could be made to improve the aesthetic and functional qualities of the models.

Realistic renderings have also been important to communicate models to NZALS staff, and the NZALS CEO to show tangible ideas and what they would look like if manufactured. These have been created in Keyshot.

continued...

Physical prototypes

Many 3D sketch models have been produced with FDM 3D printing (design and function based prototypes) via CAD modeling.

Printed sketch models and prototypes were reviewed by NZALS technician Aaron Johnston (and the designer), and changes were made subsequently to the CAD model if needed.

More resolved designs used Callaghan Innovations SLS 3D printing technology, to create high quality prototypes. These were similar to the envisioned SLS printing technology used at external printing provider Materialise and Shapeways. The Callaghan models would assist in resolving technical details around fit, strength, and material dimension.

3D printing at Callaghan also allowed the designer to participate in the printing process; this included understanding how print jobs were prepared, and the file management of digital models.

Critical reflection and validation

Critical reflection and validation (Atkins 2006) and expert engagement, has been an integral part of the research and design process.

A consistent and constant back and forth process with NZALS experts, was used to evaluate designs, and occurred in regular meetings, between student designer, academic supervisor and prosthetic technician.

Physical prints were discussed and next stage goals identified.

Photography was used to document models. Annotations directly onto models served as references to develop designs and identify issues. These were documented in sketchbooks to use as visual notes for later designs.

5. Interaction process

A process around how the research engaged with the NZALS, was configured after reflecting on several years of collaborative projects. To focus research, a three tiered approach was taken. The designer interacted with the prosthetic technician, who then interacted with NZALS staff and CEO.

This linear approach reduced assumptions and speculation on the designer's part, allowing for efficient feedback; effectively creating a more cohesive design team with clear goals to apply their skillsets.

Interaction and guidance with NZALS technician Aaron Johnston, and CEO Sean Grey, were paramount in providing a design and business context for the project.

Interactions with industry Professionals and technicians included communicating with 3D manufacturing technicians at Callaghan Innovation, such as Hamish Trolove.

This provided important technical knowledge on the Additive Manufacturing process, and designing the products to be Additively Manufactured.

Discussions with NZALS CEO Sean Grey, also provided important insights into the future directions of the company.

Communicating designs between the designer and CEO enabled designs to gain traction and financial backing.

In this regard, this constant interaction with NZALS technicians and industry professionals, acted as rolling, informal Discussions that lasted the length of the research.

The following diagram shows how these interactions were configured.

Figure 13. Process of interactions - Designer to NZLAS technicians and other industry (informed by previous student and NZLAS research)

6. Software

Rhino 3D modelling (CAD)

The modelling of all the parts and objects in this research were created in Rhino 6. This surface modelling software allows for the creation of complex geometries and structures with great accuracy. Nurbs surfaces or Polysurfaces, were created using curve networks, and resulted in the creation of the models that replicated the anatomy of the leg and foot.

Although these surfaces and forms resemble those found in human anatomy, Rhino's capabilities allow for the designer to refine and change more biological shapes, into product based models.

The ability to have great control over curve networks in Rhino, means small alterations in the curve network result in an updated Polysurface. This way, subtle refinements in the design of a product, can be easily made via the curve network.

Rhino 3D, also allows for these Polysurfaces to be manipulated through other surface modelling commands in the software. This is to add functional components to designs and visual details.

The software also has several commands that analyse the surfaces and surface edges of models, to ensure parts or Polysurfaces are visually uniform, structurally sound, and are able to be manufactured using 3D printing.

Netfabb

Netfabb is used in the final stages of the design process to prepare digital 3D models for 3D printing.

Netfabb has the ability to convert Polysurfaces into mesh based surfaces (a modelling type alternative to Polysurfaces that can be processed by a 3D printer). Netfabb can repair these meshes using repair scripts if any errors are identified that will not allow the part or model to be printed.

This software also allows for the polygon count of the mesh (which defines the level of detail in the model), to be reduced, while retaining the detail of the 3D model, but significantly lowering the file size.

Perforation's are also able to be added in Netfabb to make the cleaning process of the 3D model from the 3D printer more efficient.

Continued...

Due to the printing process of SLS, the powder which is sintered to create the 3D printed model will be retained in any closed cavity the model might have. Therefore it is important to add perforations to allow this powder to escape during the cleaning process.

Keyshot

Keyshot is a Digital rendering software that creates realistic digital images of what a product will look like, before it has been produced.

Keyshot provides an array of different materials and lighting environments that can be applied to a 3D model.

This way, designers can see in a simulated format, what different materials and colours will look like on their digital 3D models.

Keyshot is also useful for the visual validation of digital models, as lighting environments which are applied to the 'scene' that the model is placed into, can highlight any imperfections on the models surface.

This is comparable to looking at a dent on a car panel when the light hits it at a certain angle.

Keyshot can also be used to create realistic representation of the digital model in context, such as where it will be used and who will use it.

Figure 14. Process of digital workflow- sketch to 3D print

7. 3D printing processes

Prototyping

1. (Fused deposition modeling) FDM printing:

This form of 3D printing is an inexpensive and relatively quick way to visualise form from the digital CAD environment into physical scaled objects. FDM is a layered approach, where filament is heated and extruded through a nozzle, building the 3D form of an object through the creation of layered cross sections (Dudek 2013). This research will utilize this prototyping method, via the university's small scale FDM "UP-box" 3D printers, as well using the NZALS'S large scale FDM printer to create 1:1 (and scaled) sketch models, off which designs can be evaluated.

Refining/Proof of concept

2. Selective Laser Sintering (SLS):

Selective Laser Sintering is an AM process, that builds parts by using a laser, to selectively heat and fuse a powdered material (Stratasys 2019). Due to absence of support material, the sintering process allows for objects to be created with a much higher level of resolution and precision. (Fina at el 2017).

This printing technology is a robust process to build parts for all stages of the production lifecycle, from prototyping to low/high volume manufacturing. (Stratasys 2019)

3. Multi jet fusion printing (MJF):

Other 3D printing technologies such as selective laser sintering (SLS) and fused deposition modelling (FDM) are multiple-pass methods, which limits their build speed.

Multi Jet Fusion is a powder-based technology, but does not use lasers. This allows MJF to be able to create parts up to 10 time faster than SLS processes. A fusing agent is jetted where particles need to be selectively molten, and a detailing agent is jetted around the contours to improve part resolution. While light is passed over the surface of the powder bed, the jetted material captures the heat and helps distribute it evenly (Materialise 2020).

Furthermore, creators of the HP jet fusion 5200 printer explain how MJF produces high precision parts with a significantly higher strength than the SLS method. HP (2019)

The NZALS currently uses the MJF production method with the material PA-12, via Shapeways, to create the 3D printed sockets that may have an imminent integration into their prosthetic fabrication.

Figure 15. Benefits of 3D printing applied to the creation of the prosthetic products created in this research

8. 3D printing materials

Prototyping

ABS

ABS is a General use ‘go-to’ plastic. It will be used extensively in the printing of 3D sketch models, due to its low cost and accessibility within the university, and the NZALS workshop. It is low impact and relatively inaccurate, depending on the layer thickness it is printed at. It is not suitable for final product applications.

- Quick turn prototypes and 1:1 functional models
- Inexpensive
- Accessible
- Able to be printed in house

PA-11

PA-11 is a tensile, durable, and high impact-resistant material. PA-11 is Qualified for low to mid volume production. The NZLAS has yet to experiment with PA-11 in its prosthetic development; however the opportunity to work with Callaghan Innovation (mentioned earlier) allows access to this material and SLS printing technology within NZ. This will be used for high quality prototypes to validate 3D models.

- Excellent surface resolution/feature details
- Low moisture absorption
- High strength
- Able to be post processed with colour (dye)
- High impact resistance
- Structurally lightweight

Proof of concept/ refinement

PA-12

PA-12 is ideal for high-volume functional prototypes. It's high quality finish makes it perfect for final proof of concept models, and is appropriate for off the shelf applications. The NZALS currently utilize MJF, PA-12 for the production of their 3D printed below the knee sockets.

- Excellent surface resolution/feature details
- Low moisture absorption
- High strength
- High impact resistance
- Structurally lightweight

Ultrasint TPU

Ultrasint TPU is thermoplastic Polyurethane that is similar to hard rubber. The material is hard wearing, has good shock absorption qualities, and rubber-like elasticity. This Material is introduced later in the design process as an experimental production material for the Easylimb prosthetic. Ultrasint TPU is produced with Multi Jet Fusion technology; Ultrasint TPU has smooth surfaces and high surface detail.

- Good impact resistance
- Rubber like flexibility
- Extremely durable
- Useable in wet environments

Figure 16. Screenshot of the Materialise website, where 3D files can be uploaded and printed with selected materials and printing processes (Materialise 2020)

9. External 3D printing

The use of external 3D printing company Materialise and Shapeways, allows designers and customers to create 3D printed products through an online service, giving the customer control over the production process and localizing the manufacturing of components.

“Ultimately, we empower our customers to transition towards a digital manufacturing process and to launch innovations that have the potential to forever change the faces of their industries.”(Materialise 2020).

Shapeways (2020), discusses that the availability to 3D printing is changing the manufacturing ecosystem. This allows for the production of small-run prototype production, and small-run production cycles; without the need for complex mass manufacturing techniques for the output of large-run production.

Materialise has access to industry leading AM technology, including SLS and MJF.

Due to its accessible website, which includes pricing, technical specifications, material capabilities and limitations, receiving an off- the- shelf quality product, is a very linear process.

It requires very little expertise and organization to process a digital 3D model into production.

This research will utilize Materialise and Shapeways for external 3D printing for prototyping and the final production of concepts.

These high quality, validating concepts, will be paramount in showing the NZALS staff, CEO and ultimately the amputee client, the capabilities of Additive Manufacture in the design of prosthetic products.

This will also act as an example of how an external 3D printing company could be used to streamline the production of prosthetics into the NZALS. This ties in with what is mentioned in the NZLAS statement of intent, that mentions the need for...

“Collaborating with other technology developers and users, including staff exchanges, career development, and better technology solutions for patients” (NZALS 2017).

This section applies the methodologies explained previously for the development of a range of 3D printed prosthetic products, which includes the Easycover and Easylimb.

The design process used throughout this section of the research communicates a traditional industrial design practice, and uses research through design to iterate and develop the designer's ideas into physical and refined products.

This section is an extensive and in-depth portfolio showing all design thinking and design development for both the Easycover and Easylimb.

LOGO DESIGN

This section focuses on the creation of a brand identity for the NZLAS and the products that will be created throughout this research. This includes creating symbol and text based logos that will be designed into the products created throughout the research.

Logos

Creating a brand for the NZALS, and subsequently the products designed in this research, would first mean designing a logo which would appear on the products created.

Logo concepts were developed based on relevant imagery and ideas relating to prosthetics and the existing NZLAS brand.

In the early stages of the research, the brand name was decided on to be the NZALS AM (Additive Manufacture) range.

In later stages of the research, it was decided to expand the brand name into individual product names.

This included the brand names **Easycover** and **Easylimb** to refer to the Prosthetic Fairing, and full length below-the-knee prosthetic created in this research.

Figure 17. Screenshot of the NZALS website (NZALS 2020)

Sketch ideation of logo designs. Relevant symbols and text were used to provide a context for the logo and the brand.

Figure 18. Sketch ideation of logos

Figure 19. Developed logo designs

Figure 20. Further developed logo deigns

Figure 21. Product names/ typography based logos

Figure 22. Final logo designs used throughout the research

Figure 23. Final product names/ typography based logos

EAS4COVER

This section shows the development of the first product created through this research; A fully 3D printed cosmetic Prosthetic Fairing (The Easycover). The development of this product was undertaken at different stages of the research, therefore the process shown is a mixed chronological order.

Figure 24. Screenshots taken from instagram showing existing Prosthetic Fairing designs from the companies Uynq and Alleles (Instagram 2020)

Prosthetic Fairing as an initial design challenge

A Prosthetic Fairing can be defined as a cosmetic prosthetic component that fits over an existing below-the-knee prosthetic.

These 'covers' aim to provide amputees with an anatomically accurate form, which can be used to disguise, or cover their prosthetic and its componentry. Usually, these covers are complemented by graphics, colour or pattern to meet the style or taste of the amputee.

Fairings were chosen as a starting point for this research, as they are an external component of the prosthetic. This means the product (fairing) has no clinical implications on the function of the prosthetic.

The design of an external cosmetic component, such as a fairing, was important to validate the linear process of designer and technician interaction. Designer and amputee interaction would be unnecessary, due to there being little functional impact on the prosthetic itself.

Therefore, all design reflection and evaluation would be informed by NZALS technicians through informal discussion.

(This process would continue throughout the research as it allowed for a more cohesive design team, with the designer and technician receiving clear goals that applied to their skillsets)

Fairing's were also important as an initial design challenge, as the predominantly cosmetic nature of the component/product provided a stepping off point in which to integrate branding into the product and communicate a visually coherent product range.

The processes used to create the Fairing (Easycover) would be translated into the creation of the full length prosthetic (Easylimb), which is shown further into the research.

Design brief provided by NZLAS technicians

The following is a quote provided by NZLAS technician Aaron Johnston, in context to this research and the creation of a Prosthetic Fairing.

“Traditional lower limb prostheses are often complemented by the fabrication of a soft foam cover, which gives the form of a user's leg, based on measurements from the contralateral side (side of the amputees existing anatomical leg)

“Although these foam covers are desirable for amputees to feel more confident and complete as an individual, they are only sufficient for a short period of time due to the nature of flexing, moving and contact with other surfaces particularly when sitting or in an active situation, or not suitable at all due to deterioration from these activities. Furthermore these covers cannot be removed once fixed to the prosthetic”

“In light of these observations, the need for a more user friendly and robust cosmetic cover design is required. Especially for typically active users. The cosmetic cover or Fairing, should incorporate the user's personal needs and aesthetic, ease of use and fitting for clinicians, technicians and the uses.

Researching and evaluating existing faring designs that also use additive manufacturing like the UNYQ product range could be a solution to these challenges” (Aaron Johnston 2020)

Design challenges identified though discussion with NZALS technicians

- Fit suitably around the given prosthetic socket to ensure a snug and secure fit.
- Allow NZALS technicians to gain access to components on the prosthetic, including grub screws and component adaptors.
- Ideally, include some sort of shock absorption in the design to maintain the structural integrity of the product itself.
- Be as minimal in componentry as possible.
- Be able to be designed for, and produced with AM

Design and branding considerations defined by the designer

- Include NZLAS branding in the product.
- Resemble the anatomical form of an existing biological limb.
- Allow for the design to be expressed in a visually coherent product range, which addresses amputees' styles, taste and visual preferences.
- Produce a final prosthetic fairing that resembles the expectation of an off-the-shelf product. This product should appeal to 90% of the amputee clientele.

Figure 25. Physical and digital Prosthetic references

Prosthetic reference

A physical example of an existing below the knee prosthetic was provided to the designer by NZLAS technicians, to inform the design of the Fairing (Easycover).

This was complemented by a digital 3D Rhino file of the prosthetic to use as reference in the digital CAD space.

These prosthetic references allowed the designer to examine existing componentry, and create accurate 3D models based off existing measurements and dimensions.

This prosthetic reference would be used throughout the research as a guiding physical and digital precedent to inform 3D models and prototypes.

Figure 26. Measurements from the prosthetic reference

Taking measurements from the traditional prosthetic to inform the design of the Easycover

Figure 27. Anatomical reference and evaluation #1

Figure 28. Anatomical reference and evaluation #2

Anatomical reference

An accurate interpretation of anatomical form was needed to guide the overall form of the Easycover. To translate this anatomical form into the 3D space, a 3D anatomical model was introduced into Rhino to use as a reference, to create the generic calf shape of the Easycover.

This anatomically accurate model was scaled accordingly in a mesh based program (Mesh-Mixer), to align with the existing 3D model of the prosthetic (provided by the NZLAS), and align to its proportions.

Important dimensions were identified from this 3D model, which included pin pointing the widest and narrowest points on the anterior and posterior profiles of the calf, as well as on the exterior and interior profiles of the calf.

From this anatomical model, the initial form of the Easycover was created using a curve cage in Rhino, where the defining curves would allow a surface to be created.

These curves that defined the surface of the model were constantly iterated to achieve the desired form of the Easycover.

100 101

The digital 3D prosthetic reference provided the existing prosthetic's height, alignment and the 3D printed socket's dimensions.

A curve cage was constructed based on measurements obtained from the anatomical model (see previous pages), to create a generic calf shaped form. Keeping the form of the Easycover minimal, and close to an anatomical shape, allowed the design to be applicable and desirable to a large percentage of amputees and maintained the design would be understandable and credible.

This also permitted that the Easycover could be subject to customisation in later design stages, through colour and pattern application, due to its minimal overall form.

The curve cage was constructed around the periphery of the prosthetic socket with a 4mm tolerance to allow for some additional room. This would create an easier fit of the Easycover onto the prosthetic itself, once surfaces and wall thickness had been modeled.

Constructing and manipulating the curve cage in the front and side profile views within Rhino was used to then create the 3D Polysurface of the Easycover.

The top curve on the circumference of the socket was offset by 4mm to allow for wall thickness and additional modeling. This was also to allow for a little bit of give room between the fairing and the socket

Images showing designer and technician interaction. This professional review was used to inform and analysis FMD prototypes and 3D models in Rhino.

front profile showing dimensions

side profile showing dimensions

The refined shape of the Easycover fitted to the prosthetic, with dimensions. This refined form was informed by anatomical reference and measurements from the traditional prosthetic, along with constant iteration of the curve cage in Rhino.

Figure 31. Designer and Technician interaction and Easycover form

Secure fitting of the Easycover to the prosthetic using a foam insert

Top of the Easycover fitting snugly to the socket

Figure 32. FDM printed prototype used to validate the overall form of the Easycover and its two-part design, as well as the fitting process of the Easycover to the prosthetic

Figure 33. Two-part split of the Easycover- Sketches, CAD model and Physical prototype

Two part design

One of the considerations for the design of the Easycover, was to allow NZLAS technicians to have access to components, once it had been fitted to the prosthetic.

Existing designs such as the Fairing's offered by the company Uynq, allowed little accessibility to components due to the way they are secured to the prosthetic itself. These existing designs use external components such as metal clamps to secure the fairing to the prosthetic.

One way to resolve this accessibility issue, was to split the design of the Easycover into two parts. It was decided that a removable back within the Easycover's design, would allow technicians to have access to components such as grub screws and adaptors, without having to remove the fairing entirely.

This design would also allow the amputee to access important functional components such as the suction valve on their prosthetic, without having to remove the Fairing entirely.

Figure 34. Back cover removed from the Easycover prototype

Figure 35. Foam insert used to fit the Easycover to the prosthetic

Shock absorption and secure fit

Constant discussion with NZALS technician Aaron Johnston addressed the challenge of fitting the Fairing to the prosthetic itself.

This was considered throughout the design process, however was addressed in its entirety after the initial form of the Fairing was developed and refined.

One of the main issues identified with existing commercially available Fairings, was the weakness they had with adsorbing impacts if bumped into obstacles. This directly correlates to the way these Fairings are secured to the prosthetic itself.

Often existing non-3D printed Fairings (constructed out of hard plastic), would crack or shatter when subject to impact. This was identified due to these products being secured by way of clamps around the metal pylon of the prosthetic. These securing points would transfer the energy of the impacts into the brittle plastic and cause it to weaken and crack.

It was not a favorable idea to add external componentry to the Easycover, as this would add to the complexity of the design and undermine the usability of the product.

Solution

The solution to this design problem, was in the form of a foam insert. This insert would fit tightly around the existing prosthetic components including the metal pylon.

Foam inserts are readily available at NZLAS centers, and take under 5 minutes to make by an NZLAS technician. These inserts sufficiently hold the Easycover securely onto the prosthetic, while also providing a buffer between the Fairing and the prosthetic components.

This means any impacts would be absorbed by the foam and significantly reduce the risk of the Easycover being compromised.

It also minimises the clinical involvement of NZLAS technicians fitting the Fairing to the prosthetic.

Figure 36. Final modelling process - curve cage to polysurface #1

Figure 37. Final modelling process - curve cage to polysurface #2

Top of the Easycover showing the fitment against the prosthetic socket. This also shows the two-part design of the Easycover, allowing access to components on the prosthetic. The logo insert is also shown.

Figure 38. Top of the Easycover Prosthetic Fairing

Figure 39. Final CAD model of the Easycover fitted to the prosthetic model

Figure 40. Ghosted views of the Easycover and prosthetic on an anatomical model

Figure 41. Fairing from the amputee's perspective

Prosthetic

Allen key

Foam insert

Front of the Easycover

Back of the Easycover

Figure 42. Components and tools needed to fit the Easycover to the prosthetic

Validating prototype

The refined digital 3D model of the Easycover was sent to external 3D printing company Shapeways.

The NZALS has used Shapeways as an external printing provider to produce their high quality MJF printed sockets. As the quality of this printing service was already validated, it was decided to send the 3D files of the Easycover to Shapeways for printing.

The refined prototype was printed in PA-11 using SLS printing technology. This high quality print would validate if the CAD modelling of the Easycover had been successful.

It would also provide a physical model resembling that of an off-the-shelf product.

No post processing would be needed to fit the Easycover to the prosthetic.

Materialise would also be used later in the design process to print validating prototypes.

Figure 43. Shapeways website home page

Shapeways website

1.

2.

3.

4.

5.

6.

7.

8.

Figure 44. Images illustrating the fitting process of the Easycover to the prosthetic #1

Figure 45. Images illustrating the fitting process of the Easycover to the prosthetic #2

Figure 46. Final validating prototype of the Easycover Prosthetic Fairing - removing back cover

Figure 47. Final validating prototype of the Easycover Prosthetic Fairing #1

Figure 48. Final validating prototype of the Easycover Prosthetic Fairing #2

Figure 49. Final validating prototype of the Easycover Prosthetic Fairing - back cover removed

Figure 50. Final validating prototype of the Easycover Prosthetic Fairing - Details and fitment

Figure 51. Final validating prototype of the Easycover Prosthetic Fairing #3

Figure 52. Final validating prototype of the Easycover Prosthetic Fairing #4

Figure 53. Final validating prototype of the Easycover Prosthetic Fairing #5

Figure 54. Example of the Easycover being worn #1

Figure 55. Example of the Easycover being worn #2

Figure 56. 3D model of the Easycover uploaded to 3D printing providers Shapeways and Materialise - With pricing (shapeways, Materialise 2020)

Easycover Pricing

Shown to the left is a pricing comparison from 3D printing providers Materialise and Shapeways. Shapeways was used at this stage of the research to print the validating prototype of the Easycover. Further on in the research, Materialise was also used to compare print quality and price. As shown, Materialise offered a cheaper alternative (\$926 NZD) to production than Shapeways (\$ 990), with the same SLS process and same material (PA-12). This could be due to the Shapeways services being marketed at producing smaller volume prints for personal use whereas Materialise is a more industry focused company and is more suited to high volume production for commercial applications.

Pricing comparison

Production cost of the Easycover printed for this research via Materialise: \$926 NZD (including tax and logistics)

Cost of a foam Fairing fabricated by an NZLAS technician: \$550-\$650 NZD

Cost of a ‘one size fits all’ Fairing from Alleles (Prosthetic fairing manufacturer): \$690-1054 NZD (excluding logistics and tax)

As shown, the price of the Easycover positions itself in the higher end of the price range, compared to existing products from the existing Prosthetic Fairing manufacturer Alleles. This excludes a mark-up from the NZLAS (Which prices the Easycover to the amputee at around \$ 1180).

However, the functional benefits of the Easycover may provide more value for money when compared to these ‘one size fits all’ designs, and also from typical foam Fairing’s created at the NZLAS.

Typically (as identified though this research) existing Fairing designs often use hard plastic in their construction. This material is often prone to cracking and degrading when knocked into objects by the amputee. These hard plastics also often result in a bulky overall construction of the Fairing with little give or room for flexibility. Along with this, existing fairing designs, such as the Alleles products, are often difficult to secure to the prosthetic, due to awkward fitting straps and clips used as the sole securing method.

The Easycover uses 3D scanning to create a Fairing that is specific to the individual. This means the product fits perfectly around the circumference of the amputee’s socket. The fitting system, via the use of a foam insert, and the products split design, make the fitting of the Easycover to the prosthetic potentially more efficient and less difficult than existing designs.

Overall, the Easycover may offer a more resolved and considered approach to a Fairing design that is highly functional and also addresses the cosmetic application. Assuming the Easycover would be worn on a day-to-day basis, its cost per wear and functional features may justify its price in comparison with similarly priced, less functional products on the market.

This is also relevant compared to the foam Fairings created at the NZALS, as these cosmetic additions tend to only last a couple of months due to wear and tear on the foam material itself.

Figure 57. Allbirds product range used as a precedent (Allbirds 2020)

Customisation/ product range

To communicate a product range for the Easycover, renders were produced in Keyshot to show product variation.

Visual mapping of existing products from Allbirds and Nike ID were used to show how one product could be expressed in a product range with different visual qualities, including colour and pattern to address an individual's design preference.

Complementary models of the Easycover, which included patterned models, were created in order to show what customisation could extend to in an envisioned marketable product range.

It is important that amputees are shown a set range of products with visual variation to appeal to different customer tastes. Much like a footwear producer e.g.: Allbirds creates different colour ways and variations of their products in order to cater for a wide customer audience.

Figure 58. Allbirds wool runners

Figure 59. Easycover product variation expressed in a product range #1

Figure 60. Easycover product variation expressed in a product range #2

Figure 61. Easycover product variation expressed in a product range - Patterned models #1

Figure 62. Easycover product variation expressed in a product range - Patterned models #1

Digital buying platforms are a standard marketing tool for most companies selling products. Streamlining AM products into the NZLAS means addressing the exposure of the products to the amputee client, in a way that realises the purchasing of a prosthetic can be a normalised shopping experience.

Typically, product variation is limited within prosthetic design and a creating a product range that can be viewed and normalised for amputees through an online platform, is a way of exposing the importance of AM products.

136 137

Figure 64. Web page – Easycover

Figure 65. Web page – Easycover – Online shop

Figure 66. Web page – Easycover – Product range

Figure 67. Web page – Easycover – Product selection

Figure 68. Web page – Easycover – Product customisation

Figure 69. Web page – Easycover – Product purchase and ordering

EASYLIMB

This section develops on the Easycover shown previously by moving toward the goal of creating a fully 3D printed below-the-knee prosthetic, (Easylimb)

The following large section of the research is broken down into the following areas where design development could be focused.

- 3D printed secondary limb (An explanation)
- Envisioned Implementation
- Initial modeling
- Socket modeling
- Sole modeling
- Final modeling

3D printed secondary limb

The development of a full 3D printed lower limb prosthesis (Easylimb), aims to build on the Easycover that has already been created in this research. This stage of the research will build on the premise of creating branded designed products for the NZALS.

This fully 3D printed lower limb (Easylimb), will provide a branded design output specific to the NZALS, that addresses the integration of an Industry 4.0 model in its design. The combination of AM or 'Smart Manufacturing' technologies within the production of the design, means amputees are able to individualize the product to meet their individual specifications and the quality that is desired and respectful to them.

With this in mind, the aim is to design a product that appeals to a wide range of amputees (and that can be tailored to their specifications), rather than being designed as a bespoke one-off product for one individual.

This prosthesis aims to integrate the scanned prosthetic socket of an individual and combine it with additional CAD modeling. The goal is to produce a full-length, lower-limb prosthesis which can be produced with Additive Manufacture, as one singular model with no external componentry, using Muti-jet Fusion or selective laser sintering technologies.

"Designed from existing diagnostic limb data and refined digitally, a fully 3D printed prosthetic aims to incorporate the comfort and structural benefits of HP Multi jet fusion Nylon printing to produce a truly all in one daily activity limb" (Aaron Johnston 2020)

Figure 70. Diagram showing the difference between a traditional prosthetic and the Easylimb (secondary prosthetic)

Figure 71. Diagram showing the limitations of a typical hand fabricated prosthetic

Secondary limbs and traditional prosthesis

Traditional lower-limb prostheses are often limited to their use in environments, and therefore limit the activities amputees can be involved in. The components used in traditional limb making are often expensive and relatively delicate when exposed to certain elements, such as water and abrasive elements, such as dirt and sand. This is certainly the case with carbon fibre flex blades used on traditional prosthesis, which are quickly prone to de-lamination if not properly maintained.

Product aims

The aim for the Easylimb is to address amputee accessibility through a secondary prosthetic limb. The Easylimb aims to be low-maintenance, user friendly and care free to use in everyday activities.

The amputee will still have access to a primary, highly functional traditional prosthetic created by NZALS technicians (potentially with the MJF 3D printed socket), but will also have the option to have access to this daily activity limb.

NZALS Technician Aaron Johnston explains the benefits of the Easylimb in the following:

“The Easylimb is beneficial to user groups that are looking for a modern and stylish below knee exoskeleton limb for everyday activity. Key aspects that make this great for the wider patient groups are this limb has fewer limitations, less ambiguity, fuss and parts.” (Aaron Johnston 2020)

Envisioned implementation

Envisioned high volume production

To bring the Easylimb to a stage where it could be a marketed high volume product, would mean developing a parametric system or software in which an amputee's diagnostic data could be translated into a master 3D model.

Parametric software can be explained as being a digital tailor. Amputees come in all different shapes and sizes, and parametric software has the ability to adapt to those shapes and sizes and tailor the 3D print to the individual. Variations of the Easylimb could be created relevant to the individual amputee's diagnostic data and socket scan.

This would allow the Easylimb to be produced in high volumes, yet still allow each individual Easylimb to be tailored specifically to the individual.

This envisioned process extends beyond this research, **as first the creation of a tangible prototype would need to be validated.**

Hence, the focus for this section of the research. The value of parametric software is explained in the future outlooks of this research.

Incorporating diagnostic data

Incorporating Digital Manufacture and Industry 4.0 production methods into the NZALS prosthetic production, allows for increased accuracy of prosthetic fitting.

This also allows products to be tailored to individual specifications, using the amputee's diagnostic data. The NZALS uses 3D scanning as a digital technology within their services to streamline the production of their MJF 3D printed sockets for amputees.

The envisioned fitting process for the Easylimb was created by leveraging off the use of this digital tool and from information gained when modelling the Easylimb (shown later).

Continued...

To allow the Easylimb to be specific to an amputees socket shape, anatomical form, and meet the amputees functional requirements, the following diagnostic data may be needed.

3D socket scan

A 3D scan of the amputee's stump would provide an accurate 3D mesh of the individuals socket shape. This would be uploaded into a digital 3D modeling software and integrated into the master 3D model (or parametric software) of the Easylimb. This would allow the product to be specifically tailored to the individual's socket shape, and ensure an accurate fit.

Shoe size and foot measurements

These may be needed to make the anatomical form of the Easylimb, proportionate to the amputees' existing biological foot. This may also allow the Easylimb to fit inside the existing footwear of the individual amputee, to improve the functionality of the product.

Weight and height

The amputee's weight may be needed to inform the structural integrity of the product. If the amputee is heavier, the model will need to have an increased wall thickness and bracing to meet functional requirements. The amputee's height may be needed (along with the existing prosthetic height), in order to make the model proportionate and match the height of the existing biological limb.

Reference image

A reference image or 3D scan of the amputee's existing biological limb could be used as reference in order to create the overall form of the Easylimb, to match closely to the amputee's existing anatomical form and create a proportionate model.

The following diagrams show a simplified and understandable process of collecting diagnostic data, specific to the amputee, and the envisioned process of the production and modeling of the Easylimb. (Figures 72,73,74)

An envisioned website design is also included, showing an example of how a digital database could be created though the collection of the amputee's diagnostic data (which would allow the amputee to view their data online, including socket scans). (Figures 75-78)

It is envisioned this online access would streamline the purchasing process of the products created in this research, as this data would be provided to the NZLAS, along with the chosen product that the amputee has purchased online.

Figure 72. Diagnostic data to 3D model

Figure 73. Production process- Diagnostic data to finished delivered product

Figure 74. Modeling process- 3D socket scan to 3D print

Figure 75. Web page – Digital database/ Online platform for the amputee to view their dignostic data

Figure 76. Web page – Amputee's account creation

Figure 77. Web page – Amputee's account with diagnostic data and personal information

Figure 78. Web page – Amputee's socket scans and diagnostic data

Figure 79. Initial design and technical considerations for the creation of the Easylimb- Mindmap

Design considerations

The diagrams shown illustrate the technical and design considerations that must be addressed in the development of the Easylimb. (Figure 79)

This includes how the product will look visually, how it can be marketed, how the product will be constructed and how the product will be used by the amputee. The most important design considerations were identified through designer and NZLAS technician interaction. These are shown to the right. (Figure 80)

A diagram of the modeling and production considerations is also shown (see figure 81)

Figure 80. Refined technical and design considerations for the creation of the Easylimb

Figure 81. Proposed modeling and production questions/ considerations relating to the creation of the Easylimb - Flow diagram

Scenario's for the use of the Easylimb

Discussion was undertaken with NZLAS technicians In order to understand where the use of the Easylimb would benefit the amputee. Applicable situations include day-to-day activities. These could include: gardening, hanging out washing, getting in and out of transportation and general use around the house e.g. cooking, leisure activities. This also included situations or environments that a traditional prosthetic may not be suitable for use. These could include use in wet environments such as the beach and the shower, or places where a traditional prosthetic might be prone to damage (i.e. De-lamination of the carbon-fibre flex foot). Figure 82 shows potential areas/ places of use.

Figure 82. Context situations/ environments where the Easylimb would benefit the amputee - taken from the designers sketchbook

This section outlines the initial design process for the creation of the Easylimb. This includes the initial CAD modeling, and the creation of an initial 1:1 scale sketch model.

Sketchbook reference of the Easylimb model to use in solving modeling challenges

An initial model of the Easylimb limb design showing how it would look on the amputee

Sketches showing the initial modeling process of integrating the socket into the Easylimb model, and making the model one single construction

Initial modeling

The initial modeling process for the Easylimb was developed from the Easycover that was produced earlier. To streamline the design process, the curve cage that was used to create the Easycover was adapted to create the defining curve cage for the first Easylimb model.

This choice was made due to the fact this existing curve cage (from the Easycover), satisfactorily resembled the anatomy of the human leg (from the socket to the ankle).

The bottom of the Easylimb model, that would resemble the anatomy of the foot, was constructed and added to the existing curve cage to create the first full scale 3D model of the Easylimb.

The socket mesh which had been used as reference for the Easycover, was also used to create the integrated socket in these initial stages.

As this socket model was a mesh based surface, a Polysurface was created in Rhino from this mesh to streamline the design process and make socket Boolean operations quicker and more predictable.

First 1:1 FDM prototype

After initial modeling had been established, and a defining curve cage constructed, the first physical reference point was created in the form of a 1:1 scale FDM sketch model of the Easylimb. This model was crude in its construction, but was important to gain knowledge on the overall form and scale of the model that had been created in Rhino.

This model was shelled at a wall thickness of 4mm and contained a solid fill around the socket due to the FDM printing process. After printing, the FDM model was analyzed to see what needed to be developed to create an accurate and refined design going forward.

Annotations were made directly onto the model to use as references in design development.

Figure 83. Images taken from the initial modeling stages of the Easylimb

1:1 scale FMD prototype being printed at the NZALS center in Wellington on the in-house printer

Figure 84. Initial modeling and initial 1:1 scale FDM model

FMD 1:1 scale print based off the first curve cage created in rhino

Annotations were made directly onto the model to identify areas where changes would need to be made. This included surface continuity, overall form and wall thickness

The foot form of the model would need to be explored in depth to produce an overall visually suitable 3D model to be translated into the refined concept

Figure 85. Initial 1:1 scale FDM prototype

The full annotated model would be split into sections to focus design direction. This would include the socket, foot/sole and overall limb form

Focusing design direction

After reflecting on the first 1:1 scale prototype, It was decided that designing the model as a whole would be too complex, and would be inefficient from a modeling and prototyping perspective.

Therefore, the development of the Easylimb was split up into sections to focus on areas that required iteration.

With these sections resolved, they would be integrated into the final model of the Easylimb, and the overall form would be analyzed further.

This process included focusing on the follow sections of the Easylimb.

1. Socket integration within the Easylimb, and how this would be designed for SLS or MJF printing.
2. Creating and refining the foot section of the Easylimb, with the added challenge of integrating a movement based component to allow the amputee to walk normally.
3. Refining the overall form and surface quality of the Easylimb, for a visually coherent product with branding and detail.

Figure 86. 1:1 prototype in comparison with an existing prosthetic

Socket development and design

The integrated socket within the Easylimb needed several test prints to validate the digital 3D models that had been made in Rhino. The digital 3D model of the socket (provided by the NZLAS), that had been used as a reference while designing the Easycover, was again used as reference in the modeling of the Easylimb.

Final socket Polysurface used in the final Easylimb design. This was converted from the original socket without the plate adaptor

Original socket mesh provided by NZALS technician with plate adaptor integrated into the bottom

Figure 87. 3D socket scan with plate adaptor and modeled socket Polysurface

Polysurface accuracy

The digital 3D based model of the prosthetic socket (provided by the NZALS) was originally a mesh based surface created off a 3D scan.

To make modeling of the Easylimb more streamlined and easy to work with, this mesh surface was converted manually to a Polysurface in Rhino. Because of this, the accuracy of the socket Polysurface used, was not as accurate as that of the socket mesh.

This Polysurface was manipulated slightly in order to create a more attractive overall form for the resulting Easylimb model. This Polysurface was remodeled in the later stages of the Easylimb's design, to make the socket as accurate as possible to the mesh based surface.

Because the Easylimb was not intended for user testing or trial, the integrated socket within the Easylimb would not need to be a direct replicant of the socket mesh provided by NZLAS technicians (and was not the first priority in the overall design)

Validating the process of actually integrating the existing socket into a full limb design was the focus. This premise was also validated by NZALS staff.

This is contextual with previous research and its limitations, as past projects have focused heavily on socket fitment which has minimized the output for more resolved and tangible products.

If the Easylimb was to be implemented into commercial production, the mesh based socket could be used to create a perfect fit for the amputee, however this would require more time to develop outside of this research.

Images of the ProX 500 SLS 3D printer at Callaghan Innovation. This technology was used to print high quality prototypes in PA11 to validate the socket integration of the Easylimb

Figure 88. ProX 500 printer at Callahan Innovation

Designing for AM production

Integral with the idea that these products use AM for production, is actually designing them to use this production process.

The Additive process means the designer must incorporate certain design features for the product to be printed. Integrating the socket into the overall form of the Easylimb would require an SLS or MJF process to validate the modeling undertaken in Rhino.

Socket models created in Rhino were not able to be printed with SLS or MJF processes, and would first have to be imported into 3D printing software (Netfabb) to prepare the 3D files for these printing processes.

This included hollowing the socket cavity between the interior shell of the socket and the exterior shell of the actual Easylimb model itself. Other operations included reducing the 3D file (STL) size and creating perforations within the model for the AM process.

Industry ties with Callaghan innovation, meant that these integrated socket models could be printed off using SLS technology in PA11 nylon.

These prints were of vital importance, and were used as reference to develop the Easylimb for Additive Manufacture. They were also used to validate the modeling and file preparation of the integrated socket within Rhino and Netfabb.

Printing at Callaghan also provided important information about materials in terms of flexibility and strength of the integrated socket.

This helped determine the wall thickness and structural integrity of the models being produced.

Understanding SLS 3D printing at Callaghan Innovation

High quality SLS printed socket models were printed at Callaghan Innovation to validate 3D modelling and file preparation.

This offered the designer the opportunity to participate in the printing process, and contributed to the understanding of how print jobs were prepared. (Figure 88)

First SLS printed prototype of the integrated socket within the Easylimb. The model was printed with a wall thickness of 3mm, which was later reduced

Annotations were made on the model to reference where changes would need to be made

Red is highlighted to show where the surface of the model was holding tension. This would need to be changed to create a uniform and visually correct form

Figure 89. First SLS printed socket

First iteration

The first SLS printed socket, based on the early design stage, was adapted from the first large scale FDM printed sketch-model. This was the first example of using the 3D file preparation software Netfabb. This was used to make the digital 3D model of the Easylimb from Rhino, suitable for SLS printing.

Annotations were made on this socket print for later reference.

Red areas on the model indicate where surface tension was being held as a result of modeling in Rhino. This was resolved later on to create more uniform and visually correct surface.

This model was printed with a wall thickness of 3mm, which was deemed unnecessarily thick for the application. This was reduced in later iterations.

Figure 90. First socket iteration in Netfabb

The Second developed SLS print validated the developed modeling process of integrating the socket into the Easylimb design.

Annotations were made to identify where logo placements and details could be placed. The final developed concept would use an updated socket Polysurface to make the integrated socket more accurate to the 3D mesh socket that was provided by NZLAS technicians

The final model would use a more evenly curved top for visual continuity

Second iteration

Later in the design process, an updated socket was printed using the same SLS process from Callaghan Innovation.

This updated model included an updated and more refined external surface, which addressed the surface tension that was held in the previous model.

This iteration was also shelled to a thickness of 2mm, which proved flexible yet structurally strong.

More perforations were also added in Netfabb, for a more efficient cleaning process after printing. This also included constructing a large perforation outlet that would be hidden under a logo insert on the back of the limb.

This model acted as a final validation tool to move on to designing the resolved product and resolved integrated socket.

Figure 91. Second iteration of the SLS printed socket

Figure 92. Second Socket iteration in Netfabb

This section focuses on the bottom half of the Easylimb's design, which includes the foot section of the model. The design process within this section aims to develop and refine the overall form of the foot, and focuses on the creation of a flexible sole to be integrated into the Easylimb's overall design.

Flex of the traditional
prosthetic through the
use of a carbon fibre flex
foot

Toe

Heel

Arch

Under the weight of the amputee,
the flex foot bends enough to
allow the amputee to stride in a
similar manner to that of an
original anatomical limb

Foot modeling and sole design

Design direction shifted to focus on the bottom half of the Easylimb. This area was particularly complex in its design.

The main design aims within this section focused on achieving the following

1. To develop the foot form of the model to appear proportionate and consistent to the rest of the Easylimb model.
2. To achieve and initiate movement within the foot area of the model, similar to that in a traditional prosthetic. Existing prosthetics often use a carbon fibre blade which allows for movement and the amputee to stride in a normal manner. (Figure 93)
3. To design the foot section of the limb so that it can be worn inside an existing piece of footwear.
4. Keep the design minimal in componentry to minimize production cost and model complexity.
5. Being careful to not compromise design language by overcomplicating designs and creating complex form.

Figure 93. Carbon fibre flexfoot used on a traditional prosthetic - how movement is initiated

Figure 94. Visual mapping of footwear designs through boards on Pinterest - screenshot from the designer's collation (Pinterest 2020)

Footwear as a guiding precedent

The influence of footwear on this section of the design process was significant. Visual analysis and mapping of existing footwear designs, including sole design, 3D printed footwear and overall footwear form, were used to guide design direction both for function and form.

As there were a limited amount of tutorial videos detailing the design process of footwear in a CAD setting, visual analysis of existing designs proved vital for incorporating design subtly into the sole and foot of the Easylimb.

Images that were deemed relevant, were saved to boards on Pinterest, or cut out into sketchbooks and used as reference when designing in CAD.

Small Details that were overlooked at times in the modeling process, such as the continuity of curves cages, and the placement of design details, helped inform the designer to create refined visual form in the CAD space.

Figure 95. Sketches from the designers workbook

Figure 96. Creating the foot model of the Easylimb – informed by iteration and 3D prototyping

Foot Modeling

A first stage design goal was to first create the foot form itself. As the human anatomy of the foot is very complex, it was decided to base the initial foot model off measurements from the prosthetic that had already been used as reference throughout the research.

To create a form that would be similar to a foot but have the design language of a product, a visual analysis into the footwear industry was used.

This included visually mapping the shape and form of shoe soles, and the profile curves of existing products to achieve a surface that would appear proportionate and visually satisfactory.

Just like the 3D modeling methods used previously, a curve cage was created in a separate modeling file to create the foot shape.

This separate file could then be imported and joined to the curve cage that would be used to define the overall form of the full limb.

Figure 97. Taking measurements of the foot shell

Measurements off the traditional prosthetic foot shell to inform the modeling process

Iterations of foot form

The overall foot form was heavily iterated to create a surface that was visually satisfactory. This involved altering 3D models, and curve cages, sometimes by millimeters, to gain the desired outcome. These models were printed using FDM, to see physical representations of the digital 3D models. The images shown illustrate the iterative process of developing the foot section of the Easylimb. These models were annotated to use as references for later designs.

Figure 98. Foot iteration

Annotated iteration of an FDM foot model

ABS FDM models printed to explore the foot shape of the Easylimb

Figure 99. Foot model iterations printed using FDM 3D printing- rapid prototyping

Various pages of sketches exploring how flex could be achieved in the Easylimb, including flexible sole concepts and subtractive ways of creating flex, similar to a traditional prosthetic and its use of a carbon fibre blade

Figure 100. Sketching exploring how movement could be achieved within the Easylimb's design

Designing movement

Existing prosthetics often use a carbon fibre flex-foot to allow the amputee to stride in a normal manner similar to a biological limb.

The amputee's weight allows the carbon fibre blade to compress and flex, providing a similar amount of movement compared to a normal leg. (See Figure 93)

A range of concepts were created exploring ways in which movement could be achieved in the overall Easylimb design, similar to an existing carbon fibre flex-foot used on a traditional prosthetic.

It was decided that the design of a movement based feature for the Easylimb, would need to use as minimal componentry as possible, yet address this functional characteristic of movement.

It was important to not over complicate the Easylimb's overall envisioned 'care free' usability in relation to increasing its functionality through a movement based design feature.

The balance between function and form would have to be considered comprehensively in order to keep the overall appearance of the Easylimb as uncomplicated and visually understandable as possible.

Figure 101. Sketches form the designers workbook

Various sketches

Identified areas where movement occurs in the anatomical form of the leg and foot. This informed overall foot form, and areas where flexibility could be achieved within the Easylimb's design

Analysis of important movement points within the foot

Figure 102. Anatomical reference of the human foot - 3D based mesh

Anatomical Reference

Anatomical foot form was used to locate the positions that were relevant to initiate movement within the human foot, and that may be transferable to the Easylimb model.

It was decided that the main areas in human anatomy that allow for movement, would be located at four central spots, the toe, heel, arch and ankle.

These positions were taken into consideration throughout the ideation phase.

Figure 103. Foot curve cage in Rhino

Developed curve cage with identified areas of movement/ flexibility

Figure 106. Branding and detail within Allbirds products

Design details

The foot section of the Easylimb would be a focal point within the overall Easylimb's construction, and include some sort of a material transition (from the white SLS body of the Easylimb to the dark grey flexible TPU component of the Easylimb). Therefore, branding detail and overall visual form would have to be considered within this area to achieve visual consistency throughout the whole Easylimb model.

A visual mapping of foot wear designs, helped guide the placement of design details and branding. Allbirds was used as a standout precedent. Allbirds products were used as a guiding precedent as they are minimalistic in form, yet still contain design details that link the product to the brand and are visually credible to the consumer.

Figure 107. Flexible sole models created in this research

Design detail informed by visual analysis

Figure 108. Collation of images from the initial sole concepts

Flexible sole concepts

To achieve movement within the Easylimb's design, a flexible sole was chosen as a route to explore. These sole concepts would aim to use Ultrasint TPU in their production.

It was hypothesized, that under the weight of the amputee, a Flexible TPU sole could compress and flex, both in lateral and longitudinal directions. And, that this, could allow the amputee to stride in a manner similar to an existing prosthetic that uses a carbon fibre flex foot.

These soles would be a component of the Easylimb's design, which would be printed separately and then be fixed to the primary, rigid, PA-12 main construction of the model.

The physical properties of the TPU material were unknown in this stage of the research, excluding information about the material from the chart provided online.

A sample print of the material was needed to understand its flexible properties, however, this would mean waiting weeks for the sample prints to return from overseas, and delay the design of the soles significantly.

This prompted the creation of two flexible sole designs that would act as test prints to understand the properties of the material. These prints would also help validate the overall visual form of the sole and foot design created in the CAD 3D space.

Visualizing the 3D models as high quality physical prototypes would be important to refine overall form and details further and bring the final Easylimb to a stage where it would be a resolved tangible concept.

An educated guess was made on the initial construction of the flexible soles. This was informed by the chart provided by Materialise on their website. This included hypothesizing a suitable wall thickness for the sole models, and how they would join to the rest of the PA-12 printed model.

Figure 109. Modeling of test print No.1 in Rhino and cross section

Test print No.1

Informed by the Materialise chart, the first test print was hollowed with a wall thickness of 5mm. This model included a curved toe area, which was informed by the anatomical reference of movement points within the human foot. It was hypothesized this would provide the amputee with more movement in the toe area when stepping off when taking a stride. A simple plug system would be used to join the sole to the PA-12 main construction of the full limb, with the intent that the sole would be glued in place after being fitted.

Figure 110. Modelling of test print No.2 in Rhino and cross section

Cross section of the sole concept rendered in the TPU material

Test print No.2

The second test model followed a visual form and design similar to that of typical footwear sole design. This meant the sole would protrude slightly out from the original surface of the Easylimb model and have a flat profile. Informed by the Materialise chart, this test print was hollowed with a wall thickness of 7mm. The same plug system was used for this test print.

Figure 111. Technical drawing and dimensions of test print No.1 & test print No.2

Figure 112. Ordering and Pricing of the TPU soles via the Materialise website

First TPU test prints printed from Materialize

Figure 113. Flexible sole test prints from Materialise - Printed using Multi Jet Fusion in TPU

A simple Plug system was used to attach the TPU sole to the SLS printed main body of the Easylimb

Figure 114. Flexible sole test prints #1

Easylimb branding designed into the print

Visual analysis of footwear informed the details and overall form of the print

Figure 115. Flexible sole test prints #2

The top half of the limb was perforated to see how details would look in the MJF material

The top half of the model was also printed to see how the soles would fit to the rest of the limb

Figure 116. Flexible sole test prints #3

Top half of the model with perforation details

Sole print next to the corresponding foot section with joining plugs and corresponding joining holes

Figure 117. Flexible sole test prints #4

Swept up toe on the model

Flat sole model

Branded details including product name and logo

Figure 118. Flexible sole test prints #5

Images showing the flexibility of the TPU test prints. These prints proved to bend along the longitudinal axis, however rotational flexibility was minimal. This would be addressed in further test prints

Figure 119. Demonstration of flexibility – Initial TPU test prints

Test print findings

Test print No.1

The first model with a wall thickness of 5mm produced a satisfactory amount of flexibility for an initial test print. Its design allowed for movement in the toe, as hypothesized when compressed under a simulated weight. As an educated guess, 5mm seemed to be a wall thickness that was not overly rigid, but could also be improved by reducing wall thickness. This 5mm wall thickness provided sufficient insights into the material quality of the TPU, off which to develop the soles further. To maximize the potential of the design and material, this 5mm wall thickness would be reduced to around 2-3mm in later iterations.

Fitment

Overall, the simple plug system worked well and securely held the sole onto the rigid part of what would be the rest of the Easylimb's construction. If the sole's wall thickness was to be reduced, the plugs located in the center of the sole would have to be moved to the periphery, where the boundary of the model would be more ridged. It was found that the more flexible the sole, the more difficult it would be to push the plugs into their designated locating holes.

Visual analysis

The branded details that were designed into the sole worked very well. Embossed and debossed details were modeled at a depth and height of 1mm, which proved enough for these details to be clearly seen and identify the component as a product of the NZLAS.

Test print No.2

The second test proved to be far more ridged than the first model. It was still able to be flexed as shown in the images previous. However, it was more difficult to bend than test print No.1. This model had a wall thickness of 7mm, which was deemed too thick. This relatively stiff print did not roll as well as print No.1 when joined to its ridged upper due to its wall thickness. This test print was more difficult to test for movement as the rigidity of the sole provided little feedback on how the sole would act under weight and movement when simulated.

Visual analysis

This sole shape resembled that of existing footwear designs and showed how this visual language could be translated into the Easylimb's design. This was important because the visual language used in typical footwear design is something that amputees already deem credible, as these are tangible products. This potentially would help with portraying the Easylimb's design as a tangible, credible product if credible visual language was continued to be incorporated into the overall design of the Easylimb.

Annotations were made to the printed sole concepts as reference in order to develop the designs further. This included noting surface details, and areas that needed improving such as overall shape and form.

It was identified that the joining plugs would need to be moved in order for the soles to fit correctly to the rest of the full limb if flexibility was to increase

Design findings and considerations

The overall form of sole No.1 was identified to appear quite feminine in its design, due to its curved toe. The visual appearance of the sole resembled a form similar to a women's ballet flat style shoe. This was something that needed to be considered, as the Easylimb would need to appear gender neutral in its overall form. Appearing too feminine might discourage male amputees to use the product. Functionally, the curved toe would have more potential to enable movement under compression, however the overall visual form of the Easylimb would need to be consistent to appear credible to all amputee clients.

Wall thickness in the construction of the soles would also have to be reduced to maximize the functionality of the soles designs. This would be addressed in later developed prints. Further to this, the hollow construction of both of the test prints could be developed further to minimize material use and increase flexibility by removing material.

Figure 120. Annotated models used to inform the development of the TPU soles

Developed sole designs

After gaining information about how Ultrasint TPU could be used to create movement within the overall Easylimb design, it was decided to create another set of sole prints that would build on the previous tests. This was chosen to refine the function and visual form of these flexible soles. In order to make the design process more efficient, it was decided to create two more prints based off the same updated foot model and curve cage that would be used to create the sole.

These soles would use the same modeling process used in test print No.1. In that they would be integrated into the overall surface form of the Easylimb 3D model. These developed prints would build on the flat sole design used in test print No.2.

These new tests would aim to reduce the material used in the designs previous and improve the flexibility through reducing wall thickness. The flat sole design was chosen as a path to follow as visually, it appeared more consistent and compatible with the rest of the Easylimb's visual form.

These developed prints also aimed to explore if a flat sole design could be made as functional as a model with a larger toe area like test print No.1. It was hypothesized that these models (with a significantly thinner wall thickness) would be able to be compressed enough under the amputee's weight and allow for a comparatively normal stride.

Figure 121. Collation of images from the updated sole concepts

Fully hollow model with 2mm wall thickness

The midsection of the model was hollowed to reduce material and increase rotational flexibility

Cross section of the developed sole concepts rendered in the TPU material

Developed sole No.1

This design used the same hollow construction as the previous test models. In this case, the wall thickness was reduced to 2mm in the hope this would improve the flexibility of the sole significantly. A section on the top of the sole was removed to increase flexibility by ensuring the overall construction would not be as rigid. Guide rails were built that corresponded to this cut out section of the sole on the rigid SLS part of the Easylimb.

This was an alternative to the large locating/fitting plugs used previously, located in the centers of the models. As this model was hypothesized to be very flexible it would mean locating plugs positioned in the center of the sole would be very difficult to fit into their guide holes. Plugs were also moved to the perimeter of the sole for this reason.

Figure 122. Modelling of developed sole print No.1 in Rhino and cross section

Joining plugs moved to the periphery of the model and size reduced

Hollowed double walled construction

Figure 123. Modelling of developed sole print No.2 in Rhino and cross section

Cross section of the developed sole concepts rendered in the TPU material

Developed sole No.2

This sole print used a different overall construction with a double wall thickness around the perimeter of the sole, complemented with a hollowed out mid-section and a wall thickness of 3mm. The hope for this model is that it would reduce material and be easier to fit to the rigid part of the full limb. This would be because the locating plugs would have more of a foundation to press against when being fitted to the corresponding rigid SLS print.

Figure 124. Technical drawing and dimensions of developed sole print No.1 & No.2

Figure 125. Ordering and Pricing of the developed TPU soles via the Materialise website

Figure 126. Developed flexible TPU sole prints #1

Figure 127. Developed flexible TPU sole prints #2

Two updated sole models together

Double walled model with a sunken cavity

Completely hollow model with large guide plug in the middle

Figure 128. Developed flexible TPU sole prints #3

Product and brand name

Figure 129. Developed flexible TPU sole prints #4

The plugs were moved to the periphery of the soles to allow them to fit more effectively to the rest of the full limb. The plugs were reduced from a diameter of 10mm to 5mm

Images showing the soles plugged into the top half of the full limb

Figure 130. Developed flexible TPU sole prints - Joining plugs

(The white part of the model would continue into the rest of the full limb)

Logo placement

Figure 131. Developed flexible TPU sole prints - Corresponding SLS foot model

Images showing the prints flexing in both longitudinal and lateral/ rotational directions, substantially more than the first test prints. (These soles could be bent around on themselves)

Figure 132. Demonstration of flexibility – Developed TPU sole prints #1

Images showing the flexibility of the updated TPU test prints. The lowered wall thickness and improved design allowed the prints to flex substantially more than the first batch of prints

Figure 133. Demonstration of flexibility – Developed TPU sole prints #2

Print findings

Overall foot form

The refinement of overall foot form within the Easylimb's design (shown previously) created a much more credible and resolved appearance to the sole models/ prints. Details such as branding and tread pattern were refined and positioned to make the soles more unified with the overall visual form of the Easylimb.

Developed print No.1

The 2mm wall thickness of this model produced a significantly lighter sole in terms of weight. This print was also significantly more flexible than the previous tests allowing the model to able to be bent around on itself, both in longitudinal and lateral directions. The large guide rail through the middle of the sole worked well to fit the sole into the corresponding rigid part of the full limb.

As predicted, the very flexible nature of the sole meant that it was slightly difficult to fit to the top half of the model. Locating holes were sanded slightly for the sole plugs to fit correctly.

Developed print No.2

The double walled model also had a far greater amount of flex than the first two test prints, being able to be bent around on itself. Due to its geometry, the sole fitted to the rigid part of the model very well, without any post processing. This model could have been improved further by reducing the wall thickness from 3mm to 2mm or even 1.5mm.

Visual analysis

In terms of visual language, these updated sole designs validated the iterative process of refining design language and overall form. These developed sole prints resemble a component of a product that has been considered in its design and refined to a tangible stage. Design details such as branding, are evident and may build the case for branded NZALS (or AM) products as tangible and completed designs ready for an off-the-shelf application.

Improvements/ future directions

To fully explore the flexible soles and maximize the movement within the Easylimb's design, further design iteration would be needed. Due to time restrictions placed on this research, the flexible sole design was left at a stage where it could benefit from more development to maximize its overall functionality.

This component of the Easylimb could benefit from technical input in the form of an industry expert with engineering knowledge. This could benefit the design process by allowing the designer to understand the forces and physics involved when the weight and movement of the amputee are analyzed. An improved sole design could then be made accordingly.

For the purposes of this research, the final Easylimb concept will use developed sole No.1 in its construction. This is because this model illustrates refined visual form, and due to efficiently, has already been printed.

An alternative to the Ultrasint TPU soles printed is shown later in the research under the future outlooks in regard to the Easylimb concept.

Images showing the sole model and foot section of the model being inserted into existing footwear (the white part of the model would continue into the rest of the full limb)

Figure 134. Sole and foot section of the Easylimb being fit into existing footwear

Ways of wearing the limb

To increase the usability of the Easylimb's design, several ways of wearing the product were considered. A traditional prostheses carbon fibre blade is complemented by a protective rubber foot shell, sized relative to the amputee's existing biological foot. This foot shell is then able to be fitted inside the amputee's existing footwear. The flexible sole concepts and foot models developed can be worn in the following ways to match and build on the existing way of using a traditional prosthetic.

1. Bare construction

The flexible sole concepts that were developed, were intended to be permanently fixed to the rigid SLS printed construction of the Easylimb. The material qualities of TPU, mean these soles are very durable, water resistant.

Therefore if the amputee chooses, the Easylimb can be worn without the need for any external protection such as a rubber foot shell or piece of footwear.

2. Inside existing footwear

The foot model and subsequent soles would be based off diagnostic data from the amputee, which take into consideration the users foot and shoe size. Therefore, the Easylimb is able to fit inside the amputees existing footwear. In this case the model was designed to fit a men's shoe that was roughly a men's US size 10-10.5.

Figure 135. External sole model- tread pattern and sole profile

External sole

At this stage of the design process a third way of wearing the model was developed.

An external sole was designed to fit over the existing TPU sole that would be fixed onto the Easylimb. This 'accessory' external sole, would function as a protective slip on cover for use in different environments.

This external sole would ideally be produced in a silicone rubber, that would increase grip and prevent the sole of the Easylimb from getting dirty if used outside, if not already being used within an existing piece of footwear.

This external sole also allows for added customization to the Easylimb, as amputees would be able to choose from a range of colors to suit individual preferences.

FDM test prints in TPU were used as a prototyping format to show how this would appear and fit onto the Easylimb. Several iterations were modeled on the more resolved Easylimb design to show how different tread patterns and surface profiles could be used to add customization and be applicable to different environments.

Figure 136. External sole-FDM prototype

This section aims to condense the development of the individual areas of focus within the Easylimb's design (integrated socket, sole design, branding and overall 3D modeling), into a refined and resolved product concept. The contents shown will illustrate the final modeling of the Easylimb, as well as contextual images and digital renderings of the product. The final 3D printed concept, along with a continued online purchasing platform, will also be included.

Figure 137. Easylimb 3D model constructed in Rhino

Figure 138. Easylimb construction and material breakdown

Figure 139. Final curve cage for the Easylimb constructed in Rhino

Final modeling

To create the final Easylimb model, the curve cage which had been used to develop the foot of the model, was imported and joined to the curve cage that had been used to define the overall form of the Easylimb.

When joined, this full curve cage would define the overall form and the subsequent surface that would be created for the final Easylimb model.

To achieve a surface that was uniform and visually correct throughout the whole CAD model, this curve cage was manipulated by degrees of millimeters to produce a smooth and continuous surface without any surface tension. This process was imperative to creating a refined overall form.

The surface evaluation tool in Rhino was used throughout this process to achieve a consistent surface, where any tension within the surface could be identified and resolved.

The sole of the model (which had already been printed), would fit onto the Easylimb using Boolean operations to create the locating holes for the sole to be plugged into.

The socket would also use a Boolean operation to create the socket void of the model (like shown in previous iterations).

Later, surface details would be added to the entirety of the Easylimb to express brand, and overall visual language and unify the CAD model as one refined concept.

Figure 140. Curve cage to Polysurface modelling #1

Figure 141. Curve cage to Polysurface modelling #2

Figure 142. Integrated socket with logo insert and hand grips

Final socket iteration

The final socket iteration developed further on the previous models shown in this research. The Polysurface used for Boolean operations (to create the integrated socket shape within the Easylimb), was remodeled in Rhino. Although, as mentioned previously, the accuracy of this Polysurface was not paramount, it was decided to update this socket shape to match as closely as possible to the original mesh based socket. This was a decision on the designer's behalf to make the final concept more valid.

File preparation in Netfabb followed the same process as used previously, with the socket and Easylimb model being shelled to 2mm. The logo insert was also kept (as seen on the previous socket) with perforations hidden behind it (the logo insert) to increase the efficiency of the cleaning process.

Figure 143. Final socket vs original mesh based socket scan

Figure 144. Final socket in Netfabb

Figure 145. Visual mapping of footwear details from the designer's workbook

Details

Branding and design details were again influenced by footwear design. A visual synthesis of existing products and where branded details were placed on these products was evaluated. Applicable details were interpreted and translated into the Easylimb design.

These design details, such as the logo insert, brand name, tread pattern, details on the heel counter of the model and the debussed areas at the top of the model, were added for function and visual interest. As well as, ensuring the product looked consistent as a whole product.

The debussed areas at the top of the model were added to act as an ergonomic grip for the amputee to more easily slip in and out of the Easylimb.

Figure 146. Heel counter and detail precedents

Figure 147. Heel counter detail iteration

Figure 148. Details within the Easylimb's design

Figure 149. Easylimb construction #1

Figure 150. Easylimb construction #2

Figure 151. 3D model construction

Figure 152. TPU sole and corresponding guide holes

Figure 153. Sole details

Figure 154. External sole details

Figure 155. Final Easylimb model on an anatomical 3D model to show proportions and overall form

Figure 156. Easylimb 3D model with anatomical reference #1

Figure 157. Easylimb 3D model with anatomical reference #2

Figure 158. Easylimb 3D model from the amputee's perspective

Figure 159. Digital renderings of the Easylimb fitted to the amputee #1

Figure 160. Digital renderings of the Easylimb fitted to the amputee #2

Figure 161. Digital renderings of the Easylimb fitted to the amputee #3

Figure 162. Digital renderings of the Easylimb fitted to the amputee #4

The interface of 3D printing software Netfabb, used to prepare STL files for Additive Manufacture, including repairing STL files, perforating models and reducing overall file size to streamline processing

Netfabb

To prepare the Easylimb for printing, the Polysurface model created in Rhino needed to be imported into the 3D printing software Netfabb.

This software allowed the Polysurface to be converted into an STL based mesh, which then allowed the file to be correctly prepared for SLS or MJF printing processes.

The steps for preparing the Easylimb model for Additive Manufacturing are as follows.

1. Shell model:

Performing a shell command allows the void created between the external shell of the Easylimb itself and the internal shell of the socket to retain a wall thickness. By initiating this command this void becomes hollow, and merges the integrated socket and the rest of the Easylimb model into one full component.

2. Reduce triangles and run repair script:

Reducing the triangle count of the STL file, minimizes the file size of the print overall, and allows for more efficient processing of the file when printing. A repair script fixes any errors with the STL mesh, and ensures it is suitable for printing. This includes that the model has no open faces or naked edges.

3. Add Perforations:

Adding perforations to the model allows the SLS power used in the printing process to escape from the model, particularly within the hollow area where the socket is integrated into the Easylimb.

Without these perforations, SLS powder would be trapped within this hollowed area and would not be able to be removed when cleaning after printing.

Figure 163. Final Easylimb model being prepared for 3D printing in Netfabb

Figure 164. Cut away of the Easylimb 3D model in Netfabb showing the overall construction

Figure 165. Cut away of the TPU sole and main body of the Easylimb showing the joining system

Figure 166. scaled comparison of the Easylimb to an existing prosthetic- dimensions included

Easylimb Pricing

The following shows the price comparison of the Easylimb to a basic entry level prosthetic, and a prosthetic with higher grade components for the same application (below-the-knee amputee). The Pricing information shown, was provided by NZLAS technician Aaron Johnston via the NZLAS database. There is a large variation in pricing for prosthetic fabrication, due to the availability of a range of components and fittings. Higher grade components that provide higher functionality and durability, will subsequently be more expensive. Components (whether lower cost or higher cost) are chosen in part by the NZALS technician and the amputee, through consultation to meet their preferences and overall clinical requirements.

Pricing comparison

Cost of the Easylimb printed for this research: \$1,928 NZD (including tax and logistics)

Cost of an entry level traditional prosthetic: \$2,318 NZD

Cost of a prosthetic with higher grade components: \$ 8,919 NZD

(The pricing breakdown of the components used in both the fabrication of these prostheses is shown in the following pages.)

New Zealand Artificial Limbs Center

Materialise OnSite reference : 1115450

03 June 2020

[Order these parts](#)

Part	Technology	Material Finishing	Price per part	Qty	Total
sole 1 (repaired).stl 96.1 x 269.3 x 37.8 mm Vol.: 97.6 cm³	Multi Jet Fusion	Ultrasint TPU 90A-01 Normal	110.18 EUR	1	110.18 EUR
final model (repaired) (hollowed) (perforated).stl 128.4 x 262.6 x 510.2 mm Vol.: 486.1 cm³	Laser Sintering	PA12 Normal	1107.10 EUR	1	1107.10 EUR
Parts price			1217.28 EUR		
Online discount			-121.73 EUR		
Delivery			25.00 EUR		
Total			1120.55 EUR		

Figure 167. Ordering and printing the Easylimb via Materialise - PDF quote

As shown, the production cost for the Easylimb places it below the price of an entry level traditional prosthetic.

This excludes a mark-up from the NZLAS (Which prices the Easylimb to the amputee at around \$ 2500). However, this still places the overall price of the product to the amputee client, in the lower margin of traditional prosthesis fabrication and cost.

Assuming the Easylimb would be used as a daily activity limb, the cost per wear of the product may also make it competitively priced in relation to an entry level prosthetic.

This is taking into consideration that the Easylimb may be used in environments that a traditional prosthesis cannot, such as in the shower, at the beach, or in environments that contain abrasive materials (such as dirt, mud and sand), that might damage expensive components on a traditional prosthesis.

In this context, the Easylimb may provide the amputee with far greater accessibility to activities for a similar price.

It must also be considered, that the amputee client is paying for a product that has been designed specifically for an off-the-shelf application, and addresses their individual style, taste and desired quality. These factors are limited through traditional prosthetic fabrication.

Existing prosthetics are mainly focused on being a functional appendage, with little consideration of offering the amputee a resolved, designed product, with aesthetic appeal and an anatomical resemblance.

Affordable pricing is important if the product is to be accessible to a large percentage of the amputee clientele and be produced in high volumes.

Additive Manufacture may also be able to keep the production of the Easylimb competitively priced, as minimal to no external componentry is needed, along with no additional post production processing. This is due to the Easylimb's fully 3D printed construction. Along with this, there may be minimal physical man hours needed to produce the product (apart from printer calibration and setup).

This is also valid when Digital Manufacture is also used in production. The automation of production through digital software and technologies such as 3D scanning, may allow for a more efficient creation and production of the Easylimb (and Easycover) products in the digital space. This may cut out physical technician hours significantly in the production of these products, and subsequently, lowers the price of the prosthesis to the amputee.

Along with this, there are also unseen production costs that are involved through traditional prosthetic fabrication. These can be in the form of inefficient material use through subtractive fabrication methods (i.e. cutting components to size or sanding and manipulating fibreglass/ plaster). As AM is an additive process, no excess material is wasted in the production of the Easylimb (and Easycover) and therefore no money is lost through material wastage.

The efficiency of Additive and Digital Manufacture is explained further in the future outlooks of this research.

Pricing of an entry level prosthesis

Job Details

Patient Name: [REDACTED]

Job No: [REDACTED] Job Status: In-progress

Job Type: New Limb (Primary) Job Progress: Ready for Invoice

Job description: First leg

Amputation: Left Trans Tibial

Limb: [REDACTED]

Limb Type: [REDACTED]

Funder: [REDACTED]

Job Costs: \$2,317.83

Part #	Description	Qty	Units	Unit Price	Total	In Centre	Out Centre	Picked	Not Picked	Leadtime	P/O Status
9419835210095	BOM CHECK SOCKET - TT TRANS TIBIAL	1.00		257.32	257.32			1.00	0.00	3	Picked
9419835210620	BOM LAMINATE CARBON - TT TRANS TIBIAL	1.00		460.53	0.00			0.00	1.00	3	BOM
9419835162011	OB 5R2 LAMINATION DISK	1.00	EACH	62.75	62.75	8.00	2.00	1.00	0.00	7	Picked
9419835203738	OP TI-10TC2-DH-BLACK PYLON WITH DRAIN HO	1.00	EACH	211.85	211.85	0.00	2.00	1.00	0.00	7	Picked
9419835201437	OP TI-100C-DH TI TUBE CLAMP 30MM + DRAIN	1.00	EACH	213.09	213.09	0.00	5.00	1.00	0.00	7	Picked
9419835207606	OP TI-400B 4 HOLE PYRAMID	1.00	EACH	130.89	0.00	0.00	2.00	0.00	1.00	7	Pending P/O
9419835020373	OB 1H38=L23 S/AXIS FOOT W/HEEL 10MM	1.00	EACH	170.45	0.00	-1.00	0.00	0.00	1.00	7	Pending P/O
9419835001570	OB 2R10 ADAPTOR FOOT SIZES 22-25	1.00	EACH	205.87	0.00	7.00	2.00	0.00	1.00	7	Pending P/O
Grand Total					745.01			4.00			

Figure 168. Component pricing of an entry level prosthetic (NZLAS 2020)

Pricing of a higher grade prosthesis

Job Details

Patient Name: [REDACTED]

Job No: [REDACTED] Job Status: Invoiced

Job Type: Replacement Limb Job Progress: Complete

Job description: Seal in V PTB

Amputation: Left Trans Tibial

Limb: [REDACTED]

Limb Type: [REDACTED]

Funder: ACC

Funder PO No: [REDACTED]

Job Costs: \$8,919.04

Part #	Description	Qty	Units	Unit Price	Total	In Centre	Out Centre	Picked	Not Picked	Leadtime	P/O Status
9419835210095	BOM CHECK SOCKET - TT TRANS TIBIAL	1.00		257.32	257.32			1.00	0.00		
9419835171969	OS L-551002 ICELOCK EXPULSION VALVE - BK	1.00	EACH	163.48	163.48			1.00	0.00		
9419835222630	OS L-471332 SEAL-IN V TT 32CM	1.00	EACH	908.31	908.31			1.00	0.00		
9419835002485	OB 4R42 SOCKET ADAPTOR W/PYRAMID	1.00	EACH	103.11	103.11			1.00	0.00		
9419835165531	OB 4R91 CLAMP ADAPTOR S/S 34MM	1.00	EACH	107.83	107.83			1.00	0.00		
9419835031041	OB 2R77 TUBE ADAPTOR LONG 34MM DIA	1.00	EACH	150.74	150.74			1.00	0.00		
9419835205732	OP RUSH ROGUE FOOT AS PER FORM	1.00	EACH	5,474.94	5,474.94			1.00	0.00		
Grand Total					7,165.73			0.00			

Figure 169. Component pricing of a higher grade prosthetic (NZLAS 2020)

The following communicates the Easylimb through digital renderings created in Keyshot. This illustrates the final visual appearance of the product, along with highlighting product details. The images shown also show envisioned variations of the product, in terms of colour, in order to realize the adoption of the product into a product range.

Anatomical contextual images are also shown to illustrate the scale of the model, and how it would be worn.

Figure 170. Final Easylimb - overall form

Figure 171. Final Easylimb - overall form #2

Figure 172. Final Easylimb - overall form #3

Figure 173. Final Easylimb - overall form #4

Figure 174. Final Easylimb - overall form #5

Figure 175. Final Easylimb - TPU sole and External sole

Figure 176. Final Easylimb -TPU sole

Figure 177. Final Easylimb - Exploded view

Figure 178. Final Easylimb - External sole

Figure 179. Final Easylimb -Sole and surface details

Figure 180. Final Easylimb - Overall form #6

Figure 181. Final Easylimb -External sole detail and branding

Figure 182. Final Easylimb -TPU sole and External sole

Figure 183. Final Easylimb - Overall form #7

Figure 184. Easylimb - Anatomical models #1

Figure 185. Easylimb - Anatomical models #2

Figure 186. Easylimb - Anatomical models #3

Figure 187. Easylimb - Anatomical models #4

Figure 188. Easylimb - Anatomical models #5

Figure 189. Easylimb - Anatomical models #6

Figure 190. Images showing anatomical variation and subsequent Easylimb model variation #1

Figure 191. Images showing anatomical variation and subsequent Easylimb model variation #2

Figure 192. Variations of the Easylimb to communicate the product within a product range #1

Figure 193. Variations of the Easylimb to communicate the product within a product range #2

Figure 194. Example of an Easylimb product range #1

Figure 195. Example of an Easylimb product range #2

Figure 196. Easylimb product variation with external sole #1

Figure 197. Easylimb product variation with external sole #2

Figure 198. Easylimb product variation - Anatomical models #1

Figure 199. Easylimb product variation - Anatomical models #2

Figure 200. Easylimb product variation #1

Figure 201. Easylimb product variation #2

The following shows the 1:1 tangible prototype of the Easylimb. This model was printed at Materialise in nylon PA-12 with selective laser sintering.

Note: The final prototype does not include the external sole, as this was not printed.

Figure 204. Final Easylimb - Refined physical prototype #1

Figure 205. Final Easylimb - Refined physical prototype #2

Figure 206. Final Easylimb - Refined physical prototype #3

Figure 207. Final Easylimb - Design details and internal construction

Figure 208. Final Easylimb - Refined physical prototype #4

Figure 209. Final Easylimb - Integrated socket and branding details

Figure 210. Final Easylimb - Branding and surface details, logo insert with perforations and interior construction

Figure 211. Final Easylimb - logo insert with perforations

Figure 212. Final Easylimb - TPU sole #1

Figure 213. Final Easylimb - TPU sole #2

Figure 214. Final Easylimb - Refined physical prototype #5

Figure 215. Final Easylimb - Refined physical prototype #6

Figure 216. Final Easylimb - Refined physical prototype #7

Figure 217. Final Easylimb - Refined physical prototype #8

Figure 218. Final Easylimb - Refined physical prototype #9

Figure 219. Final Easylimb - Refined physical prototype #10

The following builds on the envisioned website design shown throughout this research, and provides a context for marketing the Easylimb through an online platform. Relevant contextual images have been incorporated into the website pages, to provide a context for the use of the Easylimb. An example of how the product could be purchased and customized is also shown.

Note: The website has been condensed into relevant pages and would need to be developed further to provide a more cohesive layout.

Figure 220. Web page – Easylimb – Men's and women's products

Figure 221. Web page – Easylimb – Ordering process and product creation

Figure 222. Web page – Easylimb – Company statement/ product aims

Figure 223. Web page – Easylimb – Product features

Figure 224. Web page – Easylimb – Product benefits

Figure 225. Web page - Easylimb - Contextual narrative #1

Figure 226. Web page - Easylimb - Contextual narrative #2

Figure 227. Web page - Easylimb - Contextual narrative #3

Figure 228. Web page - Easylimb - Contextual narrative #4

Figure 229. Web page - Easylimb - Contextual narrative #5

Figure 230. Web page - Easylimb - Contextual narrative #6

Figure 231. Web page - Easylimb - Contextual narrative #7

Figure 232. Web page – Easylimb – Additive Manufacturing benefits

Figure 233. Web page – Easylimb – Online shop

Figure 234. Web page – Easylimb – Product range

Figure 235. Web page – Easylimb – Product selection

Figure 236. Web page – Easylimb – Product customisation

Figure 237. Web page – Easylimb – Product purchase and ordering

DISCUSSION AND CONCLUSION

This section contains relevant discussion to the research undertaken. This includes reflecting on the design process, and the Aims and Objectives created throughout this research. This section also aims to discuss any limitations and future propositions for this research, and the products created within.

To Explore new technology and its potential applications to limb making

The NZALS statement of intent (2017) states “3D printing has the potential to significantly evolve the service experience of patients. However, these AM technologies need to be developed in a considered way that balances the benefits and risks of AM technology to ensure the appropriate implementation of 3D printing technology into service delivery.”

The Easycover and Easylimb products validate the creation of possible prosthetic products using AM production. Both products have been designed for Additive Manufacture, throughout all stages of their design life cycles (idea generation, development, refinement).

Material selection and 3D production processes have been investigated through the design process, to maximise the potential of the products created, and ensure they have been developed in a considered way. SLS, MJF and FDM printing processes have been used for prototyping and product refinement.

Material selection for refined concepts varies from the use of high tensile nylon (PA11, PA12) and Ultrasint TPU. The selection of appropriate AM processes (e.g.: SLS, MJF) for the applications of this research, were informed by the MJF prosthetic sockets that the NZLAS at present is integrating into their prosthetic production.

This was a stepping off point in which to integrate these high quality and refined AM production processes into this research.

The use of such AM processes and materials has allowed for greater product diversity within this research, which may lead to aesthetic choice, rapid reproduction (for prototyping), and economies in prototyping in logistics, material and time.

In addition, these emerging platforms of manufacturing may allow for amputee's sockets to be integrated into products, rather than them being one component in the construction of the prosthetic.

Building on previous research, the Easycover and Easylimb communicate how these production methods can be used beyond the prototyping phase, and into the context where AM production may lead to product diversity. AM technology may allow products to be tailored specifically to individuals while realising the product can be produced in large volumes.

Design products with the intent for Additive Manufacture

Integral with the idea that these products use AM for production, is actually designing them to use this production process.

The Additive process means the designer must incorporate certain design features for the product to be printed. This includes design considerations from the initial modelling in the CAD software, to file management and material selection. 3D CAD models must contain no naked edges and have wall thicknesses to be functional to print.

Additionally, perforations must be included within the product itself if there are any voids (hollow areas), to allow the printing process to function successfully. AM technology allows for quick turnover production, both in prototyping and product refinement.

It has been an integral part of the prototyping process of this research to realise digital 3D form in a tangible, physical state.

Build the NZLAS brand identity through the creation of designed products

In order for the Additive Manufacture aims of this research to be credible to the client and company, a primary focus was to design brand into the Easycover and Easylimb.

Brand names and product names help realise the outputs of this research as realistic products for an envisioned marketable output.

Integrating AM production into the NZLAS is approaching the realization of a company goal. However, these new production processes (AM) are yet to be fully evaluated and streamlined into their prosthetic production.

The Easycover and Easylimb may provide tangible solutions to this problem. Branding within the Easycover and Easylimb was important to allow the amputee client to align with the NZLAS brand values. This includes the company's value proposition of initiating change to product development and technology with AM to provide better outcomes for amputees.

AM may create a point of difference in prosthetic production for the NZLAS. The Easycover and Easylimb illustrate credible and tangible manifestations of this point of difference, which may now be understandable to the amputee.

The opportunity for amputees to align with the NZLAS brand and understand the benefits of Am production may be realized if branding is designed into the products in question.

Use industrial design to communicate feasible products

Refining concepts from a prototyping phase into refined products, was imperative to providing understandable and credible concepts. In all stages of this research, the consideration of overall form and visual language was iterated and refined constantly.

Mastering 3D form is integral to having precise control over form development and detail. Within the 3D space, the slightest of adjustments and manipulations can have a significant impact on the form of a model in the physical state. Surface tension within a 3D model, where surfaces appear to have kinks or are not consistent is an example of this.

Surface form was dictated by the curve cages used to define the surfaces of the Easycover and Easylimb. In CAD software, these were constantly iterated by degrees of millimetres, in order to create uniformity and consistency within the products.

Anatomical reference was also key in providing a context for the form of both the Easycover and Easylimb. To make the designs appear more like products, and ensure that they would be visually coherent through a product range, anatomical form was interpreted rather than replicated exactly.

The visual analysis of existing companies and products influenced form and details. Allbirds was used throughout this research as a guiding precedent for the visual consistency within its product range. Allbirds products illustrate minimal but considered visual form. This was important to translate into this research if the designs were to meet the expectations of off-the-shelf products.

This also ties in with product diversity, as existing product ranges influenced the creation of the product ranges in this research, allowing the products to appeal to individual amputee's taste and personal style.

Other existing precedents were analysed through visual mapping, to group relevant design details or nuances within the designs.

These were then interpreted into the design process. This also included Branding and overall design details, like the heel counter and hand grips on the Easylimb. The result was one overall form (product) that may appeal to a wide market rather than a bespoke one-off-design.

Developing design intuition through practise and iteration

This research has placed importance on the role of the designer being focused on digital and physical form to create tangible concepts. As mentioned earlier in the research, to the consumer and the company, the designer's control and management of 3D form is often unseen. However, it is these calculated decisions by the designer that ultimately make the design visually and functionally desirable.

It is important to realize that the designer must focus on the nuances of design within the design process, to allow products to be credible to the client and company. When other external responsibilities are taken on by the designer, such as user testing, this may reduce the ideation and refinement of concepts.

While wider interactions may be of benefit, in the context of this research, doing this would have shifted the focus away from mastering digital and physical form. Developing design intuition was validated through the entirety of the design process as mastering 3D form and refining designs became a daily exercise.

This research illustrated how external responsibilities outside of design development, should perhaps be undertaken by professionals, with specific expertise in that field (e.g.: NZALS technician for clinical engagement). This directly correlates to the quote given by NZALS technician Aaron Johnston, which has been used throughout this research to reinforce this premise.

'We found patients were more likely to respond and contribute to design development where ideas appear to be complete, tangible, readymade and resemble the expectation of an off-the-shelf product' (Aaron Johnston 2019 NZALS prosthetic technician)

Limitations and future speculations

User testing

Throughout this research, the goal has been to provide the NZLAS with products that illustrate the importance of design and branding.

It has also been the goal to inform the amputee client, that the NZALS is a brand focused on customer and innovation. And that they, as an amputee, can be a part of these concepts and the greater innovation of the brand.

To streamline this design process and bring ideas to the stage where concepts are tangible and refined, the clinical side of prosthetic production/fitting has not been the focus.

To realise the commercial implementation of the Easycover and Easylimb, the products would first have to undergo user testing to inform the design process further.

This would include the accuracy and fitment of the integrated socket, the alignment of the products to each individuals specifications, and the overall structural integrity of the products.

Field testing would also be needed to gain feedback on day-to day-wear, use in different environments and how the products function as a whole.

It would be envisioned this user testing would become part of the future design focus for the NZALS and its design/technician team.

Where to 3D print?

The technology and accessibility of high build volume, and high quality 3D printing is limited within New Zealand. If production were to be realised at a commercial level, the access to high quality and reliable 3D printing within NZ may be a significant step in streaming 3D printed products into the NZALS; possibly bringing the company closer to an industry 4.0 business model. This may reduce logistic time and cost, production turnover and mean the NZALS could track production more carefully.

There are currently several organisations that have access to large scale SLS printers with NZ. These include Callaghan Innovation and Massey University of Auckland. However, the main purpose of these printers is as a prototyping tool. Such organisations do not have the capacity or technology to use these printers for commercial production.

It was also found through use of national printing providers, that the printers used, were often out-dated and provided an inferior production quality to that of overseas providers such as Materialise.

This may be because large external providers operate in a commercial environment and have availability to high end printing technology.

It could also be said, that the adoption of 3D printing within New Zealand has been limited because the support networks that allow the AM process to operate efficiently (e.g. online purchasing and uploading platforms), are not yet established, contrasting the platforms offered by Shapeways and Materialise.

This leaves open the question of how can national businesses progress the additive manufacturing capability and accessibility within New Zealand?

The value of Parametric software

Parametric software would allow the products created in this research to be tailored specifically to any amputee's socket shape and prosthetic specifications.

Parametric software can be explained as being a digital tailor. Amputees come in all different shapes and sizes, and parametric software has the ability to adapt to those shapes and sizes and tailor the 3D print to the individual. A suit tailor needs several reference points in order to fit a suit to an individual. These references are marked and changes are made accordingly.

Parametric software allows the technician to undertake the same process, but within a digital environment, to make the 3D print fit suitably to any amputee.

To bring the Easycover and Easylimb to a stage where they could be marketed as high volume products, would mean developing a parametric system or software in which amputee's diagnostic data could be imputed into a master 3D model.

Variations of the size and shape of the product could then be created relevant to the individual amputee.

"The goal of parametric design, therefore, is to add the dimensions and any other specific information needed for functionality and manufacturability" (Poli 2001)

In order to create a parametric system, a software specialist would need to be introduced into the design process of both the Easycover and Easylimb.

Simple sliders could be used to increase the socket volume, or reduce or increase the height of the Easycover and Easylimb models, and tailor the print to fit suitably to any individual.

Additional CAD modelling along with this parametric system could allow the products to be defined by the amputee's socket shape, their height and alignment.

It would be envisioned this parametric software/ system would become part of the future design focus for the NZALS and its design/technician team

External 3D printing capability and efficiency

The production of high volume 3D printed products ties in with the ability to access a printer that can handle the large build volume of each print.

Throughout this research, several 3D printing providers have been evaluated for the availability of printers that can physically produce 1:1 scale prototypes of the Easycover and Easylimb. Due to the large scale of the prints (height and overall volume), the accessibility to SLS and MJF printers that have the capability to print such large objects such as the Easylimb, is limited. Materialise online was the only provider that could be identified through this research that could accommodate a print with a height over 500mm with SLS technology (Height of the Easylimb being around 510mm)

However, this printing availability was only researched through online platforms, that allow prototypes to be uploaded and printed.

The benefits of using such online platforms for Am production, include a very efficient model for pricing, specification and logistics. This digital process of ordering a print removes face to face consultation and streamlines the production process, as ordering and processing prints is autotomized.

Printing capability also asks the question of how many 1:1 scale prints can be constructed on one print build. To streamline production, it would make sense to print off multiple Easycovers or Easylimbs within one print build. This of couse, depends on the size of the amputee's socket and the length and volume of each amputee's specific product. The following also tie in with print build, printer accessibly and production efficiency.

- Logistics - If multiple products can be printed at a time, it may reduce the logistic time to turn over production and provide the product to the amputee. If prints are built singularly, it may affect the production time of all models waiting to be printed as backlogs of products may be created through restricted production time. This may also mean slower logistics in terms of delivery.

- Pricing - Increasing the number of products on one print build may lower the overall production cost, as the printing material may be used more efficiently. This may lead to reduced printing cost overall, as the supplier is not having to switch out SLS powder for each single print. This may also increase time efficiency, as there is less time spent with file management and printer calibration. Efficiency within production also may mean a lower price for the product overall, as production costs may not be compensated in the final price of the product, by increasing the retail price of the product to the amputee.

Sole alternative

One of the most challenging aspects of the Easylimb's design, was integrating a flexible sole, that may allow the amputee to walk in a functional manner. Ultrasint TPU was used as experimental Additive material, to see if a degree of flexibility could be achieved within the overall Easylimb design in the form of a flexible sole design. The hypothesis behind this flexible sole was that the TPU would compress under the amputee's weight and allow for a relatively normal stride.

The developed sole prints were eventually designed where they could bend and flex significantly, in both longitudinal and rotational directions. However, the compression of the prints was restricted due to the durability and overall strength of the material. Further development (as mentioned on page 237) would be needed to bring these TPU soles to a highly functional state.

A possible alternative to the TPU sole, could take influence from existing materials used in traditional footwear design, with the use of EVA foam. This alternative sole design could be easier and cheaper to produce in large volumes than the TPU soles tested.

EVA foam is typically used in existing footwear to create the sole or the outsole of the shoe, and is what is used in most running sneakers. EVA foam has high shock absorption qualities and is lightweight and flexible. The durability of the material also makes it suitable for all environments and heavy use.

Due to the fact EVA foam requires a mould and other production machinery to be formed into the shape required, it was not suitable for this research, as the production methods required were not suitable for prototyping, unlike the quick turnover 3D printed TPU.

In a commercial context, where the Easylimb is being produced in large volumes, this material and revised sole design could provide a more functional and cost effective solution to the prototyped TPU sole.

Figure 238. EVA foam sole modelled in Rhino

Figure 239. Internal and external buisness models

Business model

Another question for the future of these products, and the brand associated with them, is what the business model for the brand would be.

There are two possible business models (among many) in which the AM brand could be expressed by the NZLAS.

1.Internal

The AM range which would include the Easycover and Easylimb, could operate within the NZALS as a part of their normal operations. This could be more favourable in terms of managing the production of the products more carefully and aligning future NZALS goals with product development. In this context the AM products would operate under the NZALS brand.

2. External

The AM range could be viewed as a separate sub-brand or entity of the NZALS, which could operate as a branch from the NZLAS brand name. This would mean the AM range of prosthetics could retain their band identity, but still be owned and operated by the NZALS.

The benefits of the external business model, may allow for a more focused approach toward the continued design development of products and marketing of the products.

Companies that express a strong brand identity and future vision through their products and experiences, often initiate change to product development and technology to stay up to-date and maintain their position as market leaders.

In this context the NZALS may retain its position of being market leaders by initiating a paradigm shift in their prosthetic production. However, this also may allow the AM products and the Easycover and Easylimb brand names to initiate a more solidified brand identity to the amputee clientele through the different brand name.

In this business model, the interaction between the designer and the NZLAS, could mean a more focused approached to product development and marketing. Much like the methodologies of this research, the designer could have more of a focus on designing brand into products if the brand (AM range) has a broad but strong vision of what the brand and its products should express to the consumer.

Technician interaction may still be an integral part of this business model as clinical fitting would still be needed, so the AM products may still be integrated into the NZLAS prosthetic production under the existing company, but identify using a separate brand name.

Figure 240. Easycover and Easylimb details

Final reflection on the design outputs (Easycover and Easylimb)

The outputs of this research, including the Easylimb and Easycover communicate the integration of emerging platforms of manufacturing within the NZLAS' prosthetic production.

The products created may provide understandable examples of merging traditional industrial design practices, with industry 4.0 based production. Extensive iteration has been critical to bring designs from a prototyping stage, to refined physical concepts, illustrating the importance of the designer's role in product development.

With Additive Manufacture, prosthetic production within the NZLAS may no longer be limited to traditional prosthetic fabrication. The integration of these new technologies, may allow products to maintain large production volumes, while addressing the individual specifications and needs of individual amputee's.

These products may also provide a bridge between the amputee client and the NZLAS. Product diversity may be realised through Additive Manufacture, by expressing products through considered product ranges, that address the lifestyles, personal style and product quality desired by the amputee clientele.

Designing brand into products is another important premise within these outputs. Configuring designs to link with value propositions of the NZLAS, may allow the amputee client to align themselves with the brand, and understand the benefits of Additive Manufacture.

The products created in this research provide examples to the NZLAS how viable Additive Manufacture can be within its prosthetic production. It illustrates how AM production can allow amputee's diagnostic data (including the socket) to be integrated into products themselves.

This may allow for further opportunities for product development, moving away from the prosthetic socket being one component amongst many in a one-off appendage.

It also illustrates to the company that design itself is an important part of business innovation. In the context with these outputs, the designer may be the initiating individual that allows an idea to be translated into credible and validating concepts off which to develop further.

Conclusion

The Easycover and Easylimb have focused on illustrating product diversity in lower-limb prosthesis using design as a conduit to comprehend emerging platforms of manufacture as a production method. Through the creation of a product prototype, the predominantly service based company of the NZLAS has been exposed to a product focused methodology and a digital design and Additive Manufacture practice.

This thesis has strived to argue for the value of design, by considering a broad base of users, resulting in concepts that seek wider appeal, and move away from the one-off bespoke appendage, which is the precedent in the current environment. Potentially, this may be seen as the development of 3D printing into an industrialized and product focused methodology.

As a visual portfolio, this research presents the NZALS, its staff and amputee clientele with an accessible narrative of the design process. Although not formalised with user testing, the product concepts created communicate a proposition that exposes the unexplored value of a branding within prosthetics, and communicates the benefits of emerging platforms of manufacture in the production of prosthetics.

It is the hope that this research can assist to support and educate a service industry about the possibilities of design, where currently and previously, design has not been actively pursued as a professional role. This research communicates that design is a valid research method and a valid practice for a service based company to initiate change within the services it offers and aid in wider business innovation.

Proposed research question:

How can design communicate the possibilities of Additive Manufacture?

Figure 241. Proposed research question

Sources

Aaker, D. (1996). Building strong brands. London, Simon and schuster.

Baynes, S., Miller, T., Guy, B. (2016). Printable prosthetics. Victoria University of wellington. Wellington. oai:researcharchive.vuw.ac.nz:10063/6685

Bhatia, S., Sharma, S.(2014). 3D printed prosthetics roll off the presses. Chemical engineering progress, Vol.110, (28-33).

Dudek, P. (2013). FDM 3D Printing Technology in Manufacturing Composite Elements. Archives of metallurgy and materials, vol 58, (1415-1418) ; DOI: 10.2478/amm-2013-0186

Fina, A., Goyanes, A., Gaisford, S & Basit, A. (2017). Selective laser sintering (SLS) 3D printing of medicines. International journal of pharmaceutics , vol 529, (285-293).

Frayling, C. (1994). Research in art and design .London : Royal college of art research papers.

Gunal, M. M. (2019). Simulation for industry 4.0 past, present, and future. Switzerland : Springer books .

Hanington, B., Martin, B. (2012). Universal methods of design . O'reilly media comany .

Herbet, N. (2005). A pleliminary investigation into the develop- ment of 3D printing of prosthetic sockets . Journal of rehabilitation

research and development , 42(2) 141-146.

Hestad, M. (2013). Branding and product design an intergrated perspective . London : Gower pub co.

Lasi, H., Fettke, P., Kemper, H., Feld, T. & Hoffmann, M. (2014). Industry 4.0 . Business and information systems engineering , Vol 6, (239-242) . DOI: 10.1007/s12599-014-0334-4

Macdonald, E., Wicker, R. (2016). Multiprocess 3D printing for increasing component functionality. Science, Vol. 353.

Mark, M., Pearson, C.S. (2001). The hero and the outlaw: Building extraordinary brands through the power of archetypes. New york: McGraw-Hill .

Mycroft, D. (2011, June 16). Head of industrial design, Hewl- ett-Packard Company.

Nussbaum, B. (1 december, 2009). Technology vs. design- what is the source of innovation? BusinessWeek.

Pavitt, J. (2000). Brand. New. V&A publishing.

Press, M. C. (2003). The design Experience: The role of design and designers in the twenty first century. United knqdom : Ashgate.

Shuxian, Z., Shuxian, Z., Wanhua, Z., Bingheng, L. (2005). 3D reconstruction of the structure of a residual limb for customising the design of a prosthetic socket . Medical engineering and physics , Vol.27, (67-74). DOI: 10.1016/j.medengphy.2004.08.015

Stappers, P. J., Giaccardi, E. (2013). Research through design. The encyclopedia of human-computer interaction.

Tovey, M. (1998). Styling and design: Intuition and analysis in industrial design. Coventry university : Coventry school of art and design.

Toma, B. (2018). Additive technologies application in transtibial socket design. Department of Social Medicine and Epidemiology: University of Rijeka.

Tsur, A . (2018). Easy technique for radiographic evaluation of stump-socket fit in below-knee amputees. Annals of physical and rehabilitation medicine, vol 61, (377) .

Zimmerman, J., Stolterman, E., Forlizzi, J. (2010). An analysis and critique of research through design: Toward a formalization of research approach. ACM conference on designing interactive systems, (310-319). DOI: 10.1145/1858171.1858228

Web-sources

Materialise onsite. (2020). Retrieved from https://www.materialise.com/en 2020

Shapeways 3D printing. (2020). Advancing the mission of Shape- ways. Retrieved from https://www.shapeways.com/

Callaghan innovation. (2020). Industry 4.0.retrieved from https:// www.callaghaninnovation.govt.nz/

Unyq. (2020). About. Retrieved from http://unyq.com/en-us/about-unyq/

NZALS. (2020). Statement of intent 2017-2021. Retrieved from https://www.nzals.co.nz/publications/statement-of-intent/

NZLAS. (2018). Annual report 2018. Retrieved from https://www.nzals.co.nz/publications/annual-report/

Evans, K., (2019).The leg shop, What happens when you lose a limb? New Zealand geographic, retrieved from https://www.nzgeo.com/stories/the-leg-shop/ WRITTEN BY KATE EVANS

Nike., Abloh, V. (2017). TEXTBOOK. Retrieved from https://content.nike.com/content/dam/one-nike/en_us/season-2018-su/NikeLab/TEN/TEXTBOOK.pdf

Stratasys. (2020). Technologies. Retrieved from https://www.stratasysdirect.com/technologieshttps://www8.hp.com/h20195/v2/getpdf.aspx/4AA7-1533ENA.pdf

Figures list

Figure 7. Allbirds. (2020). Men's Collections. Retrieved from <https://www.allbirds.co.nz/collections/mens> screenshot by author & Allbirds. (2020). Women's Collections. Retrieved from <https://www.allbirds.co.nz/collections/womens> screenshot by author

Figure 8. UNYQ. (2020) Prosthetic covers. Retrieved from <http://unyq.com/en-us/prosthetic-covers/order/> screenshot by author & Alleles design studio. (2020). Leg covers. Retrieved from <https://alleles.ca/leg-covers/> screenshot by author

Figure 9. Abloh. (2020). "TEXTBOOK". Retrieved from https://content.nike.com/content/dam/one-nike/en_us/season-2018-su/NikeLab/TEN/TEXTBOOK.pdf

Figure 10. Specsavers. (2020). Home. Retrieved from <https://www.specsavers.co.nz/> screenshot by author & Ray ban. (2020). Home. Retrieved from <https://www.ray-ban.com/usa> screenshot by author

Figure 16. Materialise. (2020). Home. Retrieved from <https://www.materialise.com/en> screenshot by author

Figure 17. NZLAS. (2020). Home. Retrieved from <https://www.nzals.co.nz/> screenshot by author

Figure 24. Instagram. (2020). UNYQ. Retrieved from <https://www.instagram.com/unyq/?hl=en> screenshot by author & Instagram. (2020) Alleles. Retrieved from <https://www.instagram.com/alleles/?hl=en> screenshot by author

Figure 43. Shapeways. (2020). Materials. Retrieved from <https://www.shapeways.com/materials> screenshot by author

Figure 56. Shapeways. (2020). Cart. Retrieved from <https://www.shapeways.com/cart> screenshot by author & Materialise. (2020). Orders. Retrieved from <https://onsite.materialise.com/en/Account/Orders> screenshot by author

Figure 57. Allbirds. (2020). Men's collection. Retrieved from <https://www.allbirds.co.nz/collections/mens> screenshot by author

Figure 58. Allbirds. (2020). Women's wool runners. Retrieved from <https://www.allbirds.com/collections/womens-runners> screenshot by author

Other

Figure 94. Pintrest. (2020). Footwear. Retrieved from <https://www.pinterest.nz/mattmcgowan29/footwear/> screenshot by author

Figure 104. Materialise. (2020). Ultrasint TPU. Retrieved from materialise.com/en/manufacturing/materials/ultrasint-tpu-90a-01 screenshot by author

Figure 105. 3Dprintingmedia. (2020) Ultrasint TPU. Retrieved from <https://www.3dprintingmedia.network/wp-content/uploads/2019/09/ultrasint-tpu-1.jpg>

Figure 167. Materialise.(2020). Printing quote

Figure 168,169. NZALS. (2020) Prosthetic database. Screenshot by NZLAS technician Aaron Johnston

All other figures have been created by the designer or have been taken from the designer's sketchbook.

Informal Discussion

Personal communications and Discussions with NZLAS technician Aaron Johnston were made on a monthly basis over the 12 month period of this research.

Johnston, A. (2019) Personal communications, Observations and expert opinion

Johnston, A. (2020) Personal communications, Observations and expert opinion

Adding Brand

Design, Branding and Additive Manufacture in the production of tangible prosthetic products.

© MATT MCGOWAN 2020